

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
 Data File : VU043209.D
 Acq On : 19 Apr 2021 21:02
 Operator : SY/MD
 Sample : PB135814ZHE#12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB135814ZHE#12

Quant Time: Apr 20 03:00:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	118527	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	205706	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	202522	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	94900	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	102254	51.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.64%	
35) Dibromofluoromethane	5.298	113	70696	45.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.48%	
50) Toluene-d8	7.906	98	259606	47.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.92%	
62) 4-Bromofluorobenzene	10.639	95	94435	43.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.24%	
Target Compounds						
16) Acetone	2.633	43	29353	24.76	ug/l	98
18) Methyl Acetate	2.957	43	5631	2.26	ug/l #	85
20) Methylene Chloride	3.051	84	4377	3.09	ug/l	93
43) Isopropyl Acetate	5.919	43	18973	3.57	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
Data File : VU043209.D
Acq On : 19 Apr 2021 21:02
Operator : SY/MD
Sample : PB135814ZHE#12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB135814ZHE#12

Quant Time: Apr 20 03:00:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

