

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042020\
 Data File : VU037810.D
 Acq On : 20 Apr 2020 13:44
 Operator : JC/MD
 Sample : L2319-09
 Misc : 25mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AE1

Quant Time: Apr 21 09:45:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 21 01:10:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	266877	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	258204	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	129825	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	72442	3.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.80%
7) Chloroethane-d5	1.93	69	68429	4.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.40%
11) 1,1-Dichloroethene-d2	2.59	63	100194	2.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.80%#
20) 2-Butanone-d5	4.66	46	311407	56.15	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.30%
24) Chloroform-d	5.10	84	148579	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
26) 1,2-Dichloroethane-d4	5.74	65	90834	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
32) Benzene-d6	5.76	84	301251	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
36) 1,2-Dichloropropane-d6	6.72	67	106567	4.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.80%
41) Toluene-d8	7.92	98	273735	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.40%
43) trans-1,3-Dichloropropene-	8.20	79	43755	4.29	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.80%
46) 2-Hexanone-d5	8.65	63	264454	50.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.32%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	86701	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.80%
64) 1,2-Dichlorobenzene-d4	12.21	152	106002	4.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.60%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Trichlorofluoromethane	2.16	101	3375	0.112	ug/L	92
13) Acetone	2.65	43	5254	1.437	ug/L	99
19) 1,1-Dichloroethane	3.92	63	3531	0.091	ug/L #	82
21) 2-Butanone	4.75	43	11618	1.749	ug/L	99
22) cis-1,2-Dichloroethene	4.71	96	2511	0.116	ug/L	91
31) Carbon tetrachloride	5.56	117	8061	0.323	ug/L	94
34) Trichloroethene	6.57	95	27424	1.308	ug/L	97
47) Tetrachloroethene	8.57	164	27015	1.793	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042020\
Data File : VU037810.D
Acq On : 20 Apr 2020 13:44
Operator : JC/MD
Sample : L2319-09
Misc : 25mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AE1

Quant Time: Apr 21 09:45:57 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 21 01:10:29 2020
Response via : Initial Calibration

