

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042020\
 Data File : VU037811.D
 Acq On : 20 Apr 2020 14:09
 Operator : JC/MD
 Sample : L2319-10
 Misc : 25mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AE2

Quant Time: Apr 21 01:47:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 21 01:10:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	266988	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	255518	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	128105	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	70928	3.51	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.20%
7) Chloroethane-d5	1.93	69	67603	3.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.40%
11) 1,1-Dichloroethene-d2	2.59	63	97118	2.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.20%#
20) 2-Butanone-d5	4.66	46	306615	55.26	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.52%
24) Chloroform-d	5.10	84	149010	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
26) 1,2-Dichloroethane-d4	5.74	65	89057	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
32) Benzene-d6	5.76	84	304842	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.00%
36) 1,2-Dichloropropane-d6	6.72	67	105242	4.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.60%
41) Toluene-d8	7.92	98	270383	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.20%
43) trans-1,3-Dichloropropene-	8.20	79	43286	4.29	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.80%
46) 2-Hexanone-d5	8.65	63	260739	50.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.94%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	85454	4.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.40%
64) 1,2-Dichlorobenzene-d4	12.21	152	101035	4.42	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.40%


Target Compounds					Ovalue
3) Chloromethane	1.54	50	3003	0.105 ug/L	99
13) Acetone	2.65	43	6982	1.909 ug/L	95
42) Toluene	8.00	91	9746	0.110 ug/L	94
47) Tetrachloroethene	8.57	164	199345	13.371 ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042020\
Data File : VU037811.D
Acq On : 20 Apr 2020 14:09
Operator : JC/MD
Sample : L2319-10
Misc : 25mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AE2

Quant Time: Apr 21 01:47:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 21 01:10:29 2020
Response via : Initial Calibration









#3
Chloromethane
Concen: 0.105 ug/L
RT: 1.54 min Scan# 61
Delta R.T. -0.00 min
Lab File: VU037811.D
Acq: 20 Apr 2020 14:09




Instrument :
MSVOA_U
ClientSampleId :
C0AE2


Tot Ion: 50 Resp: 3003
Ion Ratio Lower Upper
50 100
52 33.4 23.0 42.6



#13
Acetone
Concen: 1.909 ug/L
RT: 2.65 min Scan# 408
Delta R.T. -0.00 min
Lab File: VU037811.D
Acq: 20 Apr 2020 14:09

Tgt Ion: 43 Resp: 6982
Ion Ratio Lower Upper
43 100
58 29.8 0.0 64.8





#42

Toluene

Concen: 0.110 ug/L

RT: 8.00 min Scan# 2070

Delta R.T. -0.00 min

Lab File: VU037811.D

Acq: 20 Apr 2020 14:09

Instrument :

MSVOA_U

ClientSampleId :


C0AE2

Tot Ion: 91 Resp: 9746

Ion Ratio Lower Upper


91 100

92 54.0 41.1 76.3





Abundance Ion 91.15 (90.85 to 91.85): VU0

Ion 92.15 (91.85 to 92.85): VU0



Time-->



#47

Tetrachloroethene

Concen: 13.371 ug/L

RT: 8.57 min Scan# 2250

Delta R.T. -0.00 min

Lab File: VU037811.D

Acq: 20 Apr 2020 14:09

Tgt Ion: 164 Resp: 199345


Ion Ratio Lower Upper

164 100

129 95.1 68.1 126.5

131 93.2 65.9 122.5

166 130.6 89.7 166.7




Abundance Ion 163.90 (163.60 to 164.60): V

Ion 128.95 (128.65 to 129.65): V

Ion 130.95 (130.65 to 131.65): V

Ion 165.90 (165.60 to 166.60): V



Time-->

