

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042020\
 Data File : VU037814.D
 Acq On : 20 Apr 2020 15:23
 Operator : JC/MD
 Sample : VIBLK
 Misc : 25mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK

Quant Time: Apr 21 10:05:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 21 01:10:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	272571	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	259542	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	129403	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	92188	4.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.40%
7) Chloroethane-d5	1.93	69	79569	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.40%
11) 1,1-Dichloroethene-d2	2.59	63	120430	3.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.80%
20) 2-Butanone-d5	4.66	46	296121	52.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.56%
24) Chloroform-d	5.10	84	152865	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.74	65	96857	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	5.76	84	347104	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
36) 1,2-Dichloropropane-d6	6.72	67	116752	5.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.40%
41) Toluene-d8	7.92	98	312120	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
43) trans-1,3-Dichloropropene-	8.20	79	47763	4.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.20%
46) 2-Hexanone-d5	8.65	63	254910	48.58	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.16%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	87116	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.80%
64) 1,2-Dichlorobenzene-d4	12.21	152	114367	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%


Target Compounds					Ovalue
16) Methylene chloride	3.08	84	1332	0.063 ug/L	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042020\
Data File : VU037814.D
Acq On : 20 Apr 2020 15:23
Operator : JC/MD
Sample : VIBLK
Misc : 25mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Quant Time: Apr 21 10:05:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 21 01:10:29 2020
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.063 ug/L
RT: 3.08 min Scan# 540
Delta R.T. -0.00 min
Lab File: VU037814.D
Acq: 20 Apr 2020 15:23

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Tot Ion: 84 Resp: 1332

Ion	Ratio	Lower	Upper
84	100		
86	59.3	44.2	82.0
49	115.5	92.4	171.6

