

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042021\
 Data File : VU043214.D
 Acq On : 20 Apr 2021 10:37
 Operator : SY/MD
 Sample : VU0420WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK133

Quant Time: Apr 21 04:59:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 21 04:57:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	82152	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.423	117	81771	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	39020	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	22227	5.32	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.40%
7) Chloroethane-d5	1.919	69	21429	5.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.00%
11) 1,1-Dichloroethene-d2	2.572	65	13845	5.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	114.60%
20) 2-Butanone-d5	4.642	46	89455	48.82	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.64%
24) Chloroform-d	5.073	84	61079	5.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.00%
26) 1,2-Dichloroethane-d4	5.713	65	38593	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.40%
32) Benzene-d6	5.736	84	111178	5.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.80%
36) 1,2-Dichloropropane-d6	6.700	67	37595	5.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	119.00%
41) Toluene-d8	7.906	98	98309	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
43) trans-1,3-Dichloroprop...	8.186	79	15259	5.47	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	109.40%
46) 2-Hexanone-d5	8.645	63	70852	50.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.90%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	31310	5.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.80%
66) 1,2-Dichlorobenzene-d4	12.201	152	39520	5.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042021\
Data File : VU043214.D
Acq On : 20 Apr 2021 10:37
Operator : SY/MD
Sample : VU0420WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK133

Quant Time: Apr 21 04:59:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 21 04:57:57 2021
Response via : Initial Calibration

