

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042121\
 Data File : VU043256.D
 Acq On : 21 Apr 2021 17:32
 Operator : SY/MD
 Sample : M2088-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4499

Quant Time: Apr 22 04:44:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 22 04:40:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	72310	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.424	117	74509	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	35745	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	17283	4.70	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.00%
7) Chloroethane-d5	1.919	69	17191	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.20%
11) 1,1-Dichloroethene-d2	2.575	65	11424	5.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	107.40%
20) 2-Butanone-d5	4.655	46	110840	68.73	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	137.46%#
24) Chloroform-d	5.073	84	54113	5.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.80%
26) 1,2-Dichloroethane-d4	5.710	65	34685	5.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.80%
32) Benzene-d6	5.739	84	94478	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.00%
36) 1,2-Dichloropropane-d6	6.700	67	33585	5.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	116.80%
41) Toluene-d8	7.906	98	80205	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
43) trans-1,3-Dichloroprop...	8.186	79	11989	4.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.40%
46) 2-Hexanone-d5	8.642	63	65208	51.46	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.92%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	28175	5.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.40%
66) 1,2-Dichlorobenzene-d4	12.198	152	33699	5.50	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.00%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.395	85	4495	0.97	ug/L	99
9) Trichlorofluoromethane	2.144	101	847	0.10	ug/L	85
25) Chloroform	5.093	83	5226	0.53	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042121\
Data File : VU043256.D
Acq On : 21 Apr 2021 17:32
Operator : SY/MD
Sample : M2088-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4499

Quant Time: Apr 22 04:44:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 22 04:40:57 2021
Response via : Initial Calibration

