

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042121\
 Data File : VU043242.D
 Acq On : 21 Apr 2021 11:53
 Operator : SY/MD
 Sample : M2057-23
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4544

Quant Time: Apr 22 04:41:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 22 04:40:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	92303	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.423	117	92627	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	44612	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	16562	3.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.60%
7) Chloroethane-d5	1.919	69	16081	3.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.80%
11) 1,1-Dichloroethene-d2	2.575	65	10376	3.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.40%
20) 2-Butanone-d5	4.652	46	93184	45.26	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.52%
24) Chloroform-d	5.073	84	47275	3.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.20%
26) 1,2-Dichloroethane-d4	5.710	65	31408	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.80%
32) Benzene-d6	5.735	84	85950	3.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.60%
36) 1,2-Dichloropropane-d6	6.700	67	28671	4.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.20%
41) Toluene-d8	7.906	98	73855	3.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.80%#
43) trans-1,3-Dichloroprop...	8.185	79	11602	3.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.40%
46) 2-Hexanone-d5	8.645	63	56946	36.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.30%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	24072	3.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	76.60%
66) 1,2-Dichlorobenzene-d4	12.201	152	31118	4.07	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.40%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042121\
Data File : VU043242.D
Acq On : 21 Apr 2021 11:53
Operator : SY/MD
Sample : M2057-23
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4544

Quant Time: Apr 22 04:41:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 22 04:40:57 2021
Response via : Initial Calibration

