

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042123\
 Data File : VU053822.D
 Acq On : 21 Apr 2023 13:11
 Operator : JC/MD
 Sample : VU0421WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0421WBL01

Quant Time: Apr 22 03:18:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U041923DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Apr 21 09:18:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.115	96	35654	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.636	95	12589	0.900	ug/l	0.00
Spiked Amount 1.000			Recovery	=	90.000%	
68) 1,2-Dichlorobenzene-d4	12.192	152	12827	0.772	ug/l	0.00
Spiked Amount 1.000			Recovery	=	77.000%	
Target Compounds						
15) Methylene Chloride	3.048	84	2782	0.188	ug/l	# 82
58) Tetrachloroethene	8.549	164	259	0.024	ug/l	# 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042123\
Data File : VU053822.D
Acq On : 21 Apr 2023 13:11
Operator : JC/MD
Sample : VU0421WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0421WBL01

Quant Time: Apr 22 03:18:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U041923DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Apr 21 09:18:21 2023
Response via : Initial Calibration

