

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042123\
 Data File : VU053827.D
 Acq On : 21 Apr 2023 16:18
 Operator : JC/MD
 Sample : 02304-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 38150-111522

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 04/28/2023
 Supervised By :Mahesh Dadoda 04/28/2023

Quant Time: Apr 22 03:20:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U041923DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Apr 21 09:18:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.112	96	35187	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.629	95	11738	0.851	ug/l	0.00
Spiked Amount 1.000			Recovery	=	85.000%	
68) 1,2-Dichlorobenzene-d4	12.192	152	13229	0.807	ug/l	0.00
Spiked Amount 1.000			Recovery	=	81.000%	
Target Compounds						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.581	101	355059	29.235	ug/l	94
9) 1,1-Dichloroethene	2.581	96	1111	0.112	ug/l	46
15) Methylene Chloride	3.044	84	17411	1.195	ug/l	89
24) tert-Butyl Alcohol	3.253	59	24811m	24.859	ug/l	
25) Methyl tert-Butyl Ether	3.362	73	329479	11.878	ug/l	95
31) Isopropyl Ether	3.986	45	1012476	31.573	ug/l	92
32) Ethyl-t-butyl ether	4.494	59	1121380	34.568	ug/l	93
33) Tert-Amyl methyl ether	5.935	73	561428	23.686	ug/l	97
73) n-propylbenzene	10.902	120	127934	9.222	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042123\
Data File : VU053827.D
Acq On : 21 Apr 2023 16:18
Operator : JC/MD
Sample : 02304-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
38150-111522

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/28/2023
Supervised By :Mahesh Dadoda 04/28/2023

Quant Time: Apr 22 03:20:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U041923DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Apr 21 09:18:21 2023
Response via : Initial Calibration

