

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042220\
 Data File : VU037851.D
 Acq On : 22 Apr 2020 18:31
 Operator : JC/MD
 Sample : L2388-09
 Misc : 25mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFSC5

Quant Time: Apr 23 09:49:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 23 01:37:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	275266	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	269414	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	130218	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	66611	3.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.00%
7) Chloroethane-d5	1.93	69	68238	3.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.60%
11) 1,1-Dichloroethene-d2	2.59	63	91079	2.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	49.20%#
20) 2-Butanone-d5	4.66	46	301535	52.71	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.42%
24) Chloroform-d	5.10	84	146800	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
26) 1,2-Dichloroethane-d4	5.74	65	92057	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
32) Benzene-d6	5.76	84	300246	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.20%
36) 1,2-Dichloropropane-d6	6.72	67	107659	4.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.80%
41) Toluene-d8	7.92	98	259971	3.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.60%
43) trans-1,3-Dichloropropene-	8.20	79	42012	3.95	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.00%
46) 2-Hexanone-d5	8.65	63	262462	48.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.36%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	86761	4.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.80%
64) 1,2-Dichlorobenzene-d4	12.21	152	104458	4.50	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042220\
Data File : VU037851.D
Acq On : 22 Apr 2020 18:31
Operator : JC/MD
Sample : L2388-09
Misc : 25mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFSC5

Quant Time: Apr 23 09:49:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 23 01:37:32 2020
Response via : Initial Calibration

