

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042321\
 Data File : VU043313.D
 Acq On : 24 Apr 2021 00:21
 Operator : SY/MD
 Sample : M2124-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4496

Manual Integrations
 APPROVED

MMDadoda
 4/26/2021 2:46:05 PM

Quant Time: Apr 26 08:37:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 24 07:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	99995	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.423	117	99486	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	45488	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	23562	4.63	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery = 92.60%			
7) Chloroethane-d5	1.919	69	21765	4.54	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery = 90.80%			
11) 1,1-Dichloroethene-d2	2.575	65	15557	5.29	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery = 105.80%			
20) 2-Butanone-d5	4.649	46	121882	54.65	ug/L	-0.01
Spiked Amount 50.000	Range 40 - 130		Recovery = 109.30%			
24) Chloroform-d	5.073	84	65478	5.06	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 101.20%			
26) 1,2-Dichloroethane-d4	5.713	65	40724	4.66	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 93.20%			
32) Benzene-d6	5.736	84	115192	4.97	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 99.40%			
36) 1,2-Dichloropropane-d6	6.697	67	39508	5.14	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery = 102.80%			
41) Toluene-d8	7.906	98	95999	3.99	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 79.80%			
43) trans-1,3-Dichloroprop...	8.186	79	14147	4.17	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery = 83.40%			
46) 2-Hexanone-d5	8.642	63	72401	42.79	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery = 85.58%			
56) 1,1,2,2-Tetrachloroeth...	10.764	84	32370	4.79	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery = 95.80%			
66) 1,2-Dichlorobenzene-d4	12.201	152	37648	4.83	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery = 96.60%			
Target Compounds						Qvalue
13) Acetone	2.662	43	2679m	1.63	ug/L	
25) Chloroform	5.099	83	5521	0.40	ug/L	98
38) Bromodichloromethane	7.118	83	1044	0.10	ug/L	# 92
47) Tetrachloroethene	8.559	164	1964	0.34	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042321\
Data File : VU043313.D
Acq On : 24 Apr 2021 00:21
Operator : SY/MD
Sample : M2124-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4496

Manual Integrations
APPROVED

MMDadoda
4/26/2021 2:46:05 PM

Quant Time: Apr 26 08:37:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 24 07:55:52 2021
Response via : Initial Calibration

