

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042321\
 Data File : VU043298.D
 Acq On : 23 Apr 2021 17:58
 Operator : SY/MD
 Sample : M2127-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E0AA6

Quant Time: Apr 24 06:34:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 24 06:31:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	105276	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.423	117	102720	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	48602	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	24026	4.49	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.80%
7) Chloroethane-d5	1.922	69	21412	4.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.00%
11) 1,1-Dichloroethene-d2	2.575	65	15822	5.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.20%
20) 2-Butanone-d5	4.652	46	132249	56.32	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.64%
24) Chloroform-d	5.073	84	65232	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.710	65	42795	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
32) Benzene-d6	5.739	84	118596	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
36) 1,2-Dichloropropane-d6	6.700	67	39352	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.20%
41) Toluene-d8	7.906	98	100966	4.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.20%
43) trans-1,3-Dichloroprop...	8.189	79	15356	4.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.60%
46) 2-Hexanone-d5	8.642	63	79241	45.36	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.72%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	33762	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.80%
66) 1,2-Dichlorobenzene-d4	12.198	152	39780	4.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.40%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042321\
Data File : VU043298.D
Acq On : 23 Apr 2021 17:58
Operator : SY/MD
Sample : M2127-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E0AA6

Quant Time: Apr 24 06:34:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 24 06:31:00 2021
Response via : Initial Calibration

