

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042419\
 Data File : VU031472.D
 Acq On : 24 Apr 2019 15:51
 Operator : JC/SP
 Sample : K2451-23
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Apr 25 06:37:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR042419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 25 06:05:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	342456	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	326759	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	125367	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	141104	5.14	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.80%
7) Chloroethane-d5	1.68	69	110581	5.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.60%
11) 1,1-Dichloroethene-d2	2.27	63	201333	3.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.00%
20) 2-Butanone-d5	4.21	46	405593	55.18	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.36%
24) Chloroform-d	4.64	84	220707	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.60%
26) 1,2-Dichloroethane-d4	5.31	65	122457	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.00%
32) Benzene-d6	5.33	84	442911	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
36) 1,2-Dichloropropane-d6	6.33	67	152782	5.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.00%
41) Toluene-d8	7.56	98	370337	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.20%
43) trans-1,3-Dichloropropene-	7.85	79	52788	4.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.40%
46) 2-Hexanone-d5	8.31	63	281331	50.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.94%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	117728	5.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	118621	5.29	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042419\
Data File : VU031472.D
Acq On : 24 Apr 2019 15:51
Operator : JC/SP
Sample : K2451-23
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Apr 25 06:37:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR042419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 25 06:05:25 2019
Response via : Initial Calibration

