

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042420\
 Data File : VU037874.D
 Acq On : 24 Apr 2020 16:25
 Operator : JC/MD
 Sample : L2395-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0D93

Quant Time: Apr 25 02:14:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 25 01:13:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	571538	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	553964	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	223505	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	234043	46.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.74%
7) Chloroethane-d5	1.93	69	188095	48.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.20%
11) 1,1-Dichloroethene-d2	2.59	63	290646	35.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.92%
21) 2-Butanone-d5	4.66	46	394403	99.68	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.68%
24) Chloroform-d	5.10	84	365317	48.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.06%
26) 1,2-Dichloroethane-d4	5.74	65	260184	49.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.80%
32) Benzene-d6	5.76	84	787338	50.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.52%
36) 1,2-Dichloropropane-d6	6.72	67	263382	50.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.54%
41) Toluene-d8	7.92	98	722184	50.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.32%
43) trans-1,3-Dichloropropene-	8.20	79	132641	49.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.36%
47) 2-Hexanone-d5	8.66	63	291907	102.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.34%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	354079	48.57	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.14%
64) 1,2-Dichlorobenzene-d4	12.20	152	225925	51.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.90%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.60	96	5331	1.536 ug/L	# 1
20) cis-1,2-Dichloroethene	4.71	96	25692	6.048 ug/L	98
34) Trichloroethene	6.57	95	34474	8.359 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042420\
Data File : VU037874.D
Acq On : 24 Apr 2020 16:25
Operator : JC/MD
Sample : L2395-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0D93

Quant Time: Apr 25 02:14:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 25 01:13:34 2020
Response via : Initial Calibration

