

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042423\
 Data File : VU053846.D
 Acq On : 24 Apr 2023 16:55
 Operator : JC/MD
 Sample : 02392-13 40X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHE33

Quant Time: Apr 25 05:03:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 25 05:01:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	149524	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	139456	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	68706	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	38772	4.257	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.200%
7) Chloroethane-d5	1.916	69	33879	4.432	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.600%
11) 1,1-Dichloroethene-d2	2.568	65	15721	4.194	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.800%
20) 2-Butanone-d5	4.632	46	77743	46.606	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.220%
24) Chloroform-d	5.060	84	76077	4.433	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
26) 1,2-Dichloroethane-d4	5.700	65	36264	4.572	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.400%
32) Benzene-d6	5.722	84	150065	4.568	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.687	67	46237	4.652	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.896	98	137357	4.480	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
43) trans-1,3-Dichloroprop...	8.176	79	16882	4.319	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.400%
46) 2-Hexanone-d5	8.632	63	63868	46.171	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.340%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	34609	4.309	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.200%
66) 1,2-Dichlorobenzene-d4	12.188	152	48511	4.402	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.000%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.575	96	1238	0.132	ug/L	# 1
34) Trichloroethene	6.542	95	9012	0.841	ug/L	94
47) Tetrachloroethene	8.549	164	74854	8.817	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042423\
Data File : VU053846.D
Acq On : 24 Apr 2023 16:55
Operator : JC/MD
Sample : 02392-13 40X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHE33

Quant Time: Apr 25 05:03:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 25 05:01:02 2023
Response via : Initial Calibration

