

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042423\
 Data File : VU053850.D
 Acq On : 24 Apr 2023 18:32
 Operator : JC/MD
 Sample : 02433-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHE39

Quant Time: Apr 25 05:04:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 25 05:01:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	139482	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	131359	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	65410	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	39822	4.687	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.800%
7) Chloroethane-d5	1.916	69	32854	4.608	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.200%
11) 1,1-Dichloroethene-d2	2.568	65	15861	4.536	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.800%
20) 2-Butanone-d5	4.632	46	75754	48.683	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.360%
24) Chloroform-d	5.060	84	74430	4.649	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.700	65	35562	4.807	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.200%
32) Benzene-d6	5.722	84	145440	4.700	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
36) 1,2-Dichloropropane-d6	6.687	67	47393	5.062	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.200%
41) Toluene-d8	7.896	98	134008	4.640	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.800%
43) trans-1,3-Dichloroprop...	8.176	79	16982	4.612	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.200%
46) 2-Hexanone-d5	8.632	63	62945	48.308	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.620%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	35480	4.690	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.800%
66) 1,2-Dichlorobenzene-d4	12.188	152	48730	4.645	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042423\
Data File : VU053850.D
Acq On : 24 Apr 2023 18:32
Operator : JC/MD
Sample : 02433-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHE39

Quant Time: Apr 25 05:04:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 25 05:01:02 2023
Response via : Initial Calibration

