

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042519\
 Data File : VU031501.D
 Acq On : 25 Apr 2019 14:19
 Operator : JC/SP
 Sample : K2536-08DL 100X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0X93DL

Quant Time: Apr 26 02:19:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR042419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 26 01:33:16 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	263946	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	274219	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	101534	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	101909	4.82	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.40%
7) Chloroethane-d5	1.68	69	84975	5.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.40%
11) 1,1-Dichloroethene-d2	2.27	63	150169	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	4.21	46	321757	56.79	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.58%
24) Chloroform-d	4.64	84	185081	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.40%
26) 1,2-Dichloroethane-d4	5.31	65	108586	6.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	121.80%
32) Benzene-d6	5.33	84	347415	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
36) 1,2-Dichloropropane-d6	6.32	67	126536	5.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.60%
41) Toluene-d8	7.56	98	265575	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.40%
43) trans-1,3-Dichloropropene-	7.85	79	38952	4.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.20%
46) 2-Hexanone-d5	8.32	63	181984	38.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.80%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	94530	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	94463	5.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.00%

Target Compounds					Ovalue
9) Trichlorofluoromethane	1.88	101	7835	0.247 ug/L	93
34) Trichloroethene	6.18	95	396211	17.742 ug/L	98
47) Tetrachloroethene	8.23	164	4287	0.275 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042519\
Data File : VU031501.D
Acq On : 25 Apr 2019 14:19
Operator : JC/SP
Sample : K2536-08DL 100X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0X93DL

Quant Time: Apr 26 02:19:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR042419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 26 01:33:16 2019
Response via : Initial Calibration

