

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042519\
 Data File : VU031494.D
 Acq On : 25 Apr 2019 10:02
 Operator : JC/SP
 Sample : VU0425WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK10

Quant Time: Apr 26 01:39:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR042419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 26 01:33:16 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	277325	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	286690	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	108847	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	113165	5.09	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.80%
7) Chloroethane-d5	1.68	69	94662	5.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.80%
11) 1,1-Dichloroethene-d2	2.27	63	171069	4.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.80%
20) 2-Butanone-d5	4.22	46	346078	58.14	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.28%
24) Chloroform-d	4.64	84	197995	5.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.40%
26) 1,2-Dichloroethane-d4	5.31	65	110497	5.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.00%
32) Benzene-d6	5.33	84	376034	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.33	67	131349	5.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.80%
41) Toluene-d8	7.56	98	300075	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.00%
43) trans-1,3-Dichloropropene-	7.85	79	42718	4.31	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.20%
46) 2-Hexanone-d5	8.32	63	206782	42.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.56%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	108956	5.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	99374	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042519\
Data File : VU031494.D
Acq On : 25 Apr 2019 10:02
Operator : JC/SP
Sample : VU0425WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK10

Quant Time: Apr 26 01:39:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR042419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 26 01:33:16 2019
Response via : Initial Calibration

