


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU042619\
Data File : VU031505.D
Acq On : 25 Apr 2019 17:06
Operator : JC/SP
Sample : BFB
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U042619DW.M
Title : METHOD 524.2 VOLATILES DRINKING WATER
Last Update : Fri Apr 26 04:42:12 2019



AutoFind: Scans 2865, 2866, 2867; Background Corrected with Scan 2851

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	97541	PASS
75	95	30	60	51.5	258154	PASS
95	95	100	100	100.0	501034	PASS
96	95	5	9	6.6	33005	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.0	375936	PASS
175	174	5	9	7.9	29594	PASS
176	174	95	101	96.5	362773	PASS
177	176	5	9	6.2	22669	PASS