

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU042621\  
 Data File : VU043334.D  
 Acq On : 26 Apr 2021 18:37  
 Operator : SY/MD  
 Sample : M2127-21  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 E0AA2

Quant Time: Apr 26 23:44:17 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR041521WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Mon Apr 26 23:40:31 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 6.259  | 114   | 104490   | 5.00     | ug/L  | # 0.00   |
| 28) Chlorobenzene-d5          | 9.423  | 117   | 104910   | 5.00     | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.819 | 152   | 47833    | 5.00     | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.604  | 65    | 25234    | 4.75     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 95.00%   |
| 7) Chloroethane-d5            | 1.919  | 69    | 22284    | 4.45     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 89.00%   |
| 11) 1,1-Dichloroethene-d2     | 2.575  | 65    | 15916    | 5.18     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 103.60%  |
| 20) 2-Butanone-d5             | 4.649  | 46    | 135508   | 58.15    | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 116.30%  |
| 24) Chloroform-d              | 5.073  | 84    | 67986    | 5.03     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 100.60%  |
| 26) 1,2-Dichloroethane-d4     | 5.713  | 65    | 44212    | 4.84     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 96.80%   |
| 32) Benzene-d6                | 5.739  | 84    | 118665   | 4.86     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 97.20%   |
| 36) 1,2-Dichloropropane-d6    | 6.700  | 67    | 40058    | 4.94     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 98.80%   |
| 41) Toluene-d8                | 7.906  | 98    | 100608   | 3.96     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 79.20%   |
| 43) trans-1,3-Dichloroprop... | 8.185  | 79    | 15130    | 4.23     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 84.60%   |
| 46) 2-Hexanone-d5             | 8.642  | 63    | 80510    | 45.12    | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 90.24%   |
| 56) 1,1,2,2-Tetrachloroeth... | 10.764 | 84    | 33593    | 4.71     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 94.20%   |
| 66) 1,2-Dichlorobenzene-d4    | 12.198 | 152   | 39449    | 4.81     | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 96.20%   |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU042621\  
Data File : VU043334.D  
Acq On : 26 Apr 2021 18:37  
Operator : SY/MD  
Sample : M2127-21  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
E0AA2

Quant Time: Apr 26 23:44:17 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR041521WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Mon Apr 26 23:40:31 2021  
Response via : Initial Calibration

