

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042623\
 Data File : VU053914.D
 Acq On : 26 Apr 2023 15:18
 Operator : JC/MD
 Sample : 02488-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHE65

Quant Time: Apr 27 05:56:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 27 05:52:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	97398	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	92532	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	44425	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	22727	3.831	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.600%	
7) Chloroethane-d5	1.912	69	21218	4.261	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	85.200%	
11) 1,1-Dichloroethene-d2	2.565	65	10209	4.181	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	83.600%	
20) 2-Butanone-d5	4.623	46	71884	66.157	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	132.320%#	
24) Chloroform-d	5.057	84	50398	4.508	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.200%	
26) 1,2-Dichloroethane-d4	5.697	65	26845	5.196	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.000%	
32) Benzene-d6	5.723	84	96065	4.407	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.200%	
36) 1,2-Dichloropropane-d6	6.684	67	31223	4.735	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.600%	
41) Toluene-d8	7.893	98	84320	4.145	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.800%	
43) trans-1,3-Dichloroprop...	8.176	79	11439	4.411	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.200%	
46) 2-Hexanone-d5	8.629	63	55849	60.848	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	121.700%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	29103	5.461	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	109.200%	
66) 1,2-Dichlorobenzene-d4	12.189	152	32476	4.558	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	91.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042623\
Data File : VU053914.D
Acq On : 26 Apr 2023 15:18
Operator : JC/MD
Sample : 02488-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHE65

Quant Time: Apr 27 05:56:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 27 05:52:18 2023
Response via : Initial Calibration

