

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042623\
 Data File : VU053916.D
 Acq On : 26 Apr 2023 16:06
 Operator : JC/MD
 Sample : 02433-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Apr 27 05:56:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 27 05:52:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	142458	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	132244	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	64661	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.598	65	33159	3.821	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.400%
7) Chloroethane-d5	1.913	69	28432	3.904	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.000%
11) 1,1-Dichloroethene-d2	2.562	65	13715	3.841	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.800%
20) 2-Butanone-d5	4.636	46	67989	42.780	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.560%
24) Chloroform-d	5.060	84	64810	3.964	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.200%
26) 1,2-Dichloroethane-d4	5.700	65	31133	4.120	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.400%
32) Benzene-d6	5.726	84	126882	4.073	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.400%
36) 1,2-Dichloropropane-d6	6.687	67	41136	4.365	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.200%
41) Toluene-d8	7.893	98	110245	3.792	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.800%
43) trans-1,3-Dichloroprop...	8.176	79	13688	3.693	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.800%
46) 2-Hexanone-d5	8.629	63	56064	42.739	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.480%
56) 1,1,2,2-Tetrachloroeth...	10.752	84	30376	3.988	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.800%
66) 1,2-Dichlorobenzene-d4	12.189	152	41716	4.022	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042623\
Data File : VU053916.D
Acq On : 26 Apr 2023 16:06
Operator : JC/MD
Sample : 02433-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Apr 27 05:56:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 27 05:52:18 2023
Response via : Initial Calibration

