

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042623\
 Data File : VU053912.D
 Acq On : 26 Apr 2023 14:31
 Operator : JC/MD
 Sample : VIBLK160
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK160

Quant Time: Apr 27 05:55:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 27 05:52:18 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	130918	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	127118	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	60583	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	33073	4.147	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.000%
7) Chloroethane-d5	1.916	69	29151	4.356	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.200%
11) 1,1-Dichloroethene-d2	2.565	65	13684	4.170	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.400%
20) 2-Butanone-d5	4.636	46	72168	49.413	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.820%
24) Chloroform-d	5.057	84	71410	4.752	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
26) 1,2-Dichloroethane-d4	5.700	65	32929	4.742	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.800%
32) Benzene-d6	5.722	84	137203	4.582	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.687	67	43197	4.768	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	7.893	98	118128	4.227	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	8.176	79	14616	4.102	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.000%
46) 2-Hexanone-d5	8.629	63	60103	47.666	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.340%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	30934	4.225	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.600%
66) 1,2-Dichlorobenzene-d4	12.188	152	43577	4.485	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.600%
Target Compounds						
16) Methylene chloride	3.041	84	13517	1.203	ug/L	92
47) Tetrachloroethene	8.552	164	1318	0.170	ug/L	83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042623\
Data File : VU053912.D
Acq On : 26 Apr 2023 14:31
Operator : JC/MD
Sample : VIBLK160
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK160

Quant Time: Apr 27 05:55:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 27 05:52:18 2023
Response via : Initial Calibration









#16
Methylene chloride
Concen: 1.203 ug/L
RT: 3.041 min Scan# 51
Delta R.T. -0.003 min
Lab File: VU053912.D
Acq: 26 Apr 2023 14:31

Instrument :
MSVOA_U
ClientSampleId :
VIBLK160

Tgt Ion: 84	Resp: 13517
Ion Ratio	Lower Upper
84 100	
86 59.7	45.0 83.6
49 97.5	74.8 138.8



#47
Tetrachloroethene
Concen: 0.170 ug/L
RT: 8.552 min Scan# 2243
Delta R.T. 0.003 min
Lab File: VU053912.D
Acq: 26 Apr 2023 14:31

Tgt Ion: 164	Resp: 1318
Ion Ratio	Lower Upper
164 100	
129 108.5	68.3 126.8
131 73.2	64.4 119.6
166 148.9	88.3 163.9

