

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042719\
 Data File : VU031534.D
 Acq On : 26 Apr 2019 15:43
 Operator : JC/SP
 Sample : K2596-11
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFHH8

Quant Time: Apr 27 01:52:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 27 00:11:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	857078	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	824283	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	371686	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	378929	52.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.66%
7) Chloroethane-d5	1.68	69	290456	52.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.80%
11) 1,1-Dichloroethene-d2	2.27	63	532276	37.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.52%
21) 2-Butanone-d5	4.18	46	624112	96.03	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.03%
24) Chloroform-d	4.64	84	592361	45.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.18%
26) 1,2-Dichloroethane-d4	5.31	65	388482	45.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.52%
32) Benzene-d6	5.34	84	1264210	49.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.90%
36) 1,2-Dichloropropane-d6	6.33	67	431727	48.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.96%
41) Toluene-d8	7.56	98	1108567	50.28	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.56%
43) trans-1,3-Dichloropropene-	7.85	79	186922	48.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.80%
47) 2-Hexanone-d5	8.31	63	435866	113.08	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.08%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	547751	45.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.66%
64) 1,2-Dichlorobenzene-d4	11.85	152	391500	50.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.18%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042719\
Data File : VU031534.D
Acq On : 26 Apr 2019 15:43
Operator : JC/SP
Sample : K2596-11
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFHH8

Quant Time: Apr 27 01:52:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 27 00:11:31 2019
Response via : Initial Calibration

