

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042719\
 Data File : VU031548.D
 Acq On : 26 Apr 2019 22:20
 Operator : JC/SP
 Sample : K2536-20 50X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0X08

Quant Time: Apr 27 03:43:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 27 03:26:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	864115	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	797678	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	316026	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	368794	50.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.98%
7) Chloroethane-d5	1.68	69	283557	50.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.48%
11) 1,1-Dichloroethene-d2	2.27	63	520632	36.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.30%
21) 2-Butanone-d5	4.18	46	672487	102.63	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.63%
24) Chloroform-d	4.64	84	609168	45.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.98%
26) 1,2-Dichloroethane-d4	5.31	65	393810	45.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.02%
32) Benzene-d6	5.34	84	1292675	52.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.54%
36) 1,2-Dichloropropane-d6	6.32	67	435634	50.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.10%
41) Toluene-d8	7.56	98	1085417	50.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.76%
43) trans-1,3-Dichloropropene-	7.85	79	179390	48.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.98%
47) 2-Hexanone-d5	8.31	63	429257	115.08	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	115.08%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	563322	48.71	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.42%
64) 1,2-Dichlorobenzene-d4	11.86	152	349888	52.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042719\
Data File : VU031548.D
Acq On : 26 Apr 2019 22:20
Operator : JC/SP
Sample : K2536-20 50X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0X08

Quant Time: Apr 27 03:43:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 27 03:26:56 2019
Response via : Initial Calibration

