

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042721\
 Data File : VU043352.D
 Acq On : 27 Apr 2021 14:27
 Operator : SY/MD
 Sample : M2127-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Apr 27 23:49:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 27 23:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	107289	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.424	117	105654	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	48762	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	22756	4.17	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.40%
7) Chloroethane-d5	1.919	69	22428	4.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.20%
11) 1,1-Dichloroethene-d2	2.575	65	15579	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.60%
20) 2-Butanone-d5	4.658	46	134350	56.15	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.30%
24) Chloroform-d	5.073	84	63344	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
26) 1,2-Dichloroethane-d4	5.713	65	44853	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
32) Benzene-d6	5.736	84	118890	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.700	67	40303	4.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.80%
41) Toluene-d8	7.906	98	99586	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.80%
43) trans-1,3-Dichloroprop...	8.186	79	14953	4.15	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.00%
46) 2-Hexanone-d5	8.645	63	78681	43.79	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.58%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	35117	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.80%
66) 1,2-Dichlorobenzene-d4	12.198	152	40685	4.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.40%
Target Compounds						Qvalue
16) Methylene chloride	3.057	84	1383	0.18	ug/L	90
25) Chloroform	5.102	83	5132	0.35	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042721\
Data File : VU043352.D
Acq On : 27 Apr 2021 14:27
Operator : SY/MD
Sample : M2127-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Apr 27 23:49:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 27 23:46:47 2021
Response via : Initial Calibration

