

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042721\
 Data File : VU043349.D
 Acq On : 27 Apr 2021 13:15
 Operator : SY/MD
 Sample : M2127-25
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E0AA8

Quant Time: Apr 27 23:49:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 27 23:46:47 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	109354	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.423	117	110401	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	50223	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	24502	4.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.20%
7) Chloroethane-d5	1.919	69	22997	4.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.80%
11) 1,1-Dichloroethene-d2	2.575	65	15429	4.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.80%
20) 2-Butanone-d5	4.649	46	131315	53.84	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.68%
24) Chloroform-d	5.073	84	64140	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
26) 1,2-Dichloroethane-d4	5.710	65	44450	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
32) Benzene-d6	5.735	84	121710	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	6.700	67	42181	4.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.00%
41) Toluene-d8	7.906	98	105248	3.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.80%
43) trans-1,3-Dichloroprop...	8.185	79	15512	4.12	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.40%
46) 2-Hexanone-d5	8.642	63	79745	42.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.94%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	34853	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.00%
66) 1,2-Dichlorobenzene-d4	12.201	152	41780	4.85	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.00%
Target Compounds						
13) Acetone	2.662	43	39839	22.23	ug/L	# 65
16) Methylene chloride	3.054	84	1319	0.17	ug/L	80
21) 2-Butanone	4.732	43	6838	2.31	ug/L	82
25) Chloroform	5.096	83	5982	0.40	ug/L	98
42) Toluene	7.976	91	5163	0.16	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042721\
Data File : VU043349.D
Acq On : 27 Apr 2021 13:15
Operator : SY/MD
Sample : M2127-25
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E0AA8

Quant Time: Apr 27 23:49:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 27 23:46:47 2021
Response via : Initial Calibration









#13
Acetone
Concen: 22.23 ug/L
RT: 2.662 min Scan# 411
Delta R.T. -0.023 min
Lab File: VU043349.D
Acq: 27 Apr 2021 13:15




Instrument :
MSVOA_U
ClientSampleId :
E0AA8


Tgt Ion: 43 Resp: 39839
Ion Ratio Lower Upper
43 100
58 26.2 0.0 25.0#



#16
Methylene chloride
Concen: 0.17 ug/L
RT: 3.054 min Scan# 533
Delta R.T. -0.000 min
Lab File: VU043349.D
Acq: 27 Apr 2021 13:15

Tgt Ion: 84 Resp: 1319
Ion Ratio Lower Upper
84 100
86 46.7 44.9 83.5
49 132.6 79.0 146.8









#21
2-Butanone
Concen: 2.31 ug/L
RT: 4.732 min Scan# 1055
Delta R.T. -0.010 min
Lab File: VU043349.D
Acq: 27 Apr 2021 13:15




Instrument :
MSVOA_U
ClientSampled :
E0AA8


Tgt Ion: 43 Resp: 6838
Ion Ratio Lower Upper
43 100
72 21.3 15.8 47.3



#25
Chloroform
Concen: 0.40 ug/L
RT: 5.096 min Scan# 1168
Delta R.T. -0.003 min
Lab File: VU043349.D
Acq: 27 Apr 2021 13:15


Tgt Ion: 83 Resp: 5982
Ion Ratio Lower Upper
83 100
85 66.9 45.6 84.8





#42
Toluene
Concen: 0.16 ug/L
RT: 7.976 min Scan# 2064
Delta R.T. -0.000 min
Lab File: VU043349.D
Acq: 27 Apr 2021 13:15

Instrument :
MSVOA_U
ClientSampleId :
E0AA8



Tgt Ion: 91 Resp: 5163
Ion Ratio Lower Upper
91 100
92 55.6 40.9 76.0

