

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042722\
 Data File : VU048317.D
 Acq On : 27 Apr 2022 16:14
 Operator : SY/MD
 Sample : N2542-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Apr 28 04:38:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 28 04:33:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	325592	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	326627	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	153051	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	146257	48.849	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	97.700%	
7) Chloroethane-d5	1.896	69	121687	57.901	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	=	115.800%	
11) 1,1-Dichloroethene-d2	2.565	63	191546	39.928	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	79.860%	
21) 2-Butanone-d5	4.620	46	244320	95.675	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	95.670%	
24) Chloroform-d	5.064	84	270692	49.845	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.680%	
26) 1,2-Dichloroethane-d4	5.700	65	175950	54.606	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	109.220%	
32) Benzene-d6	5.726	84	572744	55.599	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	111.200%	
36) 1,2-Dichloropropane-d6	6.691	67	183611	53.913	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	107.820%	
41) Toluene-d8	7.899	98	496498	53.172	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	106.340%	
43) trans-1,3-Dichloroprop...	8.179	79	77759	51.744	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	103.480%	
47) 2-Hexanone-d5	8.636	63	171722	93.457	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	93.460%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	300244	50.267	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	100.540%	
66) 1,2-Dichlorobenzene-d4	12.195	152	198822	57.748	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	115.500%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042722\
Data File : VU048317.D
Acq On : 27 Apr 2022 16:14
Operator : SY/MD
Sample : N2542-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Apr 28 04:38:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 28 04:33:44 2022
Response via : Initial Calibration

