




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042722\
Data File : VU048316.D
Acq On : 27 Apr 2022 15:50
Operator : SY/MD
Sample : N2562-08ME
Misc : 5.53g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 16 Sample Multiplier: 1

Quant Time: Apr 28 04:38:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 28 04:33:44 2022
Response via : Initial Calibration



TIC: VU048316.D\data.ms

(7) Chloroethane-d5 (S)




1.877min (-0.032) 11.20 ug/L

response 26062

Ion	Exp%	Act%
69.10	100.00	100.00
71.10	31.60	26.56
51.00	22.10	20.72
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042722\
 Data File : VU048316.D
 Acq On : 27 Apr 2022 15:50
 Operator : SY/MD
 Sample : N2562-08ME
 Misc : 5.53g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Quant Time: Apr 28 04:38:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 28 04:33:44 2022
 Response via : Initial Calibration



TIC: VU048316.D\data.ms

(7) Chloroethane-d5 (S)**1.877min (-0.032) 23.12 ug/L m****response 53812**

Ion	Exp%	Act%
69.10	100.00	100.00
71.10	31.60	12.86#
51.00	22.10	10.03#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042722\
 Data File : VU048316.D
 Acq On : 27 Apr 2022 15:50
 Operator : SY/MD
 Sample : N2562-08ME
 Misc : 5.53g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Quant Time: Apr 28 04:38:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 28 04:33:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	360645	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	364731	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	174605	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	128386	38.713	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	77.420%	
7) Chloroethane-d5	1.877	69	53812m	23.116	ug/L	-0.03
Spiked Amount 50.000	Range 70	- 130	Recovery	=	46.240%#	
11) 1,1-Dichloroethene-d2	2.533	63	167568	31.534	ug/L	-0.04
Spiked Amount 50.000	Range 60	- 125	Recovery	=	63.060%	
21) 2-Butanone-d5	4.639	46	251840	89.034	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	89.030%	
24) Chloroform-d	5.063	84	259751	43.181	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	86.360%	
26) 1,2-Dichloroethane-d4	5.700	65	174709	48.950	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	97.900%	
32) Benzene-d6	5.719	84	543788	47.273	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	94.540%	
36) 1,2-Dichloropropane-d6	6.687	67	178866	47.033	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	94.060%	
41) Toluene-d8	7.896	98	487769	46.780	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	93.560%	
43) trans-1,3-Dichloroprop...	8.182	79	77760	46.339	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	92.680%	
47) 2-Hexanone-d5	8.665	63	182939	89.161	ug/L	0.03
Spiked Amount 100.000	Range 45	- 130	Recovery	=	89.160%	
56) 1,1,2,2-Tetrachloroeth...	10.767	84	286589	42.969	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	85.940%	
66) 1,2-Dichlorobenzene-d4	12.198	152	192713	49.064	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	98.120%	
Target Compounds						
22) 2-Butanone	4.723	43	36500	13.593	ug/L	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042722\
Data File : VU048316.D
Acq On : 27 Apr 2022 15:50
Operator : SY/MD
Sample : N2562-08ME
Misc : 5.53g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 16 Sample Multiplier: 1

Quant Time: Apr 28 04:38:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 28 04:33:44 2022
Response via : Initial Calibration

