

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042722\
 Data File : VU048304.D
 Acq On : 27 Apr 2022 11:05
 Operator : SY/MD
 Sample : VU0427MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK102

Quant Time: Apr 28 04:35:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 28 04:33:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	355518	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	360519	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	162475	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	168260	51.468	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.940%
7) Chloroethane-d5	1.906	69	128883	56.163	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	112.320%
11) 1,1-Dichloroethene-d2	2.565	63	210668	40.217	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.440%
21) 2-Butanone-d5	4.629	46	259129	92.932	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.930%
24) Chloroform-d	5.066	84	294350	49.639	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.280%
26) 1,2-Dichloroethane-d4	5.703	65	186043	52.878	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.760%
32) Benzene-d6	5.729	84	621381	54.650	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.300%
36) 1,2-Dichloropropane-d6	6.693	67	199250	53.005	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.020%
41) Toluene-d8	7.899	98	549269	53.294	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.580%
43) trans-1,3-Dichloroprop...	8.182	79	83643	50.427	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.860%
47) 2-Hexanone-d5	8.639	63	173542	85.569	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.570%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	301884	45.791	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.580%
66) 1,2-Dichlorobenzene-d4	12.195	152	203893	55.786	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042722\
Data File : VU048304.D
Acq On : 27 Apr 2022 11:05
Operator : SY/MD
Sample : VU0427MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK102

Quant Time: Apr 28 04:35:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 28 04:33:44 2022
Response via : Initial Calibration

