

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042723\
 Data File : VU053927.D
 Acq On : 26 Apr 2023 21:27
 Operator : JC/MD
 Sample : VSTD0.0536
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.05036

Quant Time: Apr 27 08:06:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 27 08:06:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	14994	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	13454	0.500	ug/L	0.02
11) 1,4-Dichlorobenzene-d4	11.742	152	7937	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	602	0.055	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.661	65	563	0.050	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.639	67	598	0.044	ug/L	0.00
8) Toluene-d8	7.845	98	1222	0.049	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.679	84	595	0.050	ug/L	0.00
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.598	62	625	0.055	ug/L #	3
6) Trichloroethene	6.489	95	779	0.051	ug/L	97
9) 1,2-Dibromoethane	8.866	107	408	0.045	ug/L #	94
12) 1,2,3-Trichloropropane	10.748	75	379	0.048	ug/L	95
13) 1,2-Dibromo-3-chloropr...	12.924	75	54	0.041	ug/L #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042723\
Data File : VU053927.D
Acq On : 26 Apr 2023 21:27
Operator : JC/MD
Sample : VSTD0.0536
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.05036

Quant Time: Apr 27 08:06:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 27 08:06:36 2023
Response via : Initial Calibration

