

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042723\
 Data File : VU053929.D
 Acq On : 26 Apr 2023 22:40
 Operator : JC/MD
 Sample : VSTD0.538
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5038

Quant Time: Apr 27 08:07:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 27 08:06:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.207	114	15239	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.346	117	13498	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.742	152	7959	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	5176	0.465	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.661	65	5531	0.482	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.639	67	6909	0.511	ug/L	0.00
8) Toluene-d8	7.830	98	12275	0.488	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.679	84	5821	0.488	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.599	62	5698	0.491	ug/L	100
6) Trichloroethene	6.489	95	7860	0.514	ug/L	100
9) 1,2-Dibromoethane	8.866	107	4822	0.528	ug/L	100
12) 1,2,3-Trichloropropane	10.748	75	4059	0.508	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.924	75	676	0.511	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042723\
Data File : VU053929.D
Acq On : 26 Apr 2023 22:40
Operator : JC/MD
Sample : VSTD0.538
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5038

Quant Time: Apr 27 08:07:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 27 08:06:36 2023
Response via : Initial Calibration

