

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042723\
 Data File : VU053932.D
 Acq On : 26 Apr 2023 23:52
 Operator : JC/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VICV041

Quant Time: Apr 27 08:25:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 27 08:10:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	15038	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.346	117	13165	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.742	152	7645	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	5157	0.469	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	94.000%
4) 1,2-Dichloroethane-d4	5.661	65	5603	0.495	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
7) 1,2-Dichloropropane-d6	6.639	67	6844	0.519	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.000%
8) Toluene-d8	7.830	98	12380	0.505	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.679	84	5968	0.513	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.598	62	5279	0.461	ug/L	99
6) Trichloroethene	6.489	95	7293	0.489	ug/L	99
9) 1,2-Dibromoethane	8.866	107	4558	0.512	ug/L	99
12) 1,2,3-Trichloropropane	10.748	75	3873	0.504	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.924	75	684	0.538	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042723\
Data File : VU053932.D
Acq On : 26 Apr 2023 23:52
Operator : JC/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VICV041

Quant Time: Apr 27 08:25:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 27 08:10:40 2023
Response via : Initial Calibration

