

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042723\
 Data File : VU053935.D
 Acq On : 27 Apr 2023 11:58
 Operator : JC/MD
 Sample : VU0427WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK016

Quant Time: Apr 28 07:56:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 27 08:10:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.207	114	14547	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	12756	0.500	ug/L	0.02
11) 1,4-Dichlorobenzene-d4	11.742	152	7613	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	4695	0.442	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	88.000%		
4) 1,2-Dichloroethane-d4	5.661	65	5222	0.477	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	96.000%		
7) 1,2-Dichloropropane-d6	6.639	67	6439	0.504	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	100.000%		
8) Toluene-d8	7.830	98	11735	0.494	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	98.000%		
10) 1,1,2,2-Tetrachloroeth...	10.679	84	5645	0.501	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	100.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042723\
Data File : VU053935.D
Acq On : 27 Apr 2023 11:58
Operator : JC/MD
Sample : VU0427WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK016

Quant Time: Apr 28 07:56:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM042723.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 27 08:10:40 2023
Response via : Initial Calibration

