

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
 Data File : VU043359.D
 Acq On : 27 Apr 2021 21:47
 Operator : SY/MD
 Sample : M2133-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB01-20210420

Quant Time: Apr 28 09:12:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	108993	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	181035	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	172799	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	82176	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	96099	52.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.90%	
35) Dibromofluoromethane	5.298	113	66840	49.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.26%	
50) Toluene-d8	7.906	98	228552	47.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.96%	
62) 4-Bromofluorobenzene	10.639	95	80874	41.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.92%	
Target Compounds						
16) Acetone	2.646	43	2257	1.79	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
Data File : VU043359.D
Acq On : 27 Apr 2021 21:47
Operator : SY/MD
Sample : M2133-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB01-20210420

Quant Time: Apr 28 09:12:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

