

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
 Data File : VU043360.D
 Acq On : 27 Apr 2021 22:09
 Operator : SY/MD
 Sample : M2133-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ118I-20210420

Quant Time: Apr 28 09:12:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	109077	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	180548	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	171719	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	80996	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	95361	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
35) Dibromofluoromethane	5.298	113	66564	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
50) Toluene-d8	7.906	98	227186	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
62) 4-Bromofluorobenzene	10.639	95	79618	41.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.84%	
Target Compounds						
16) Acetone	2.649	43	3244	2.71	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
Data File : VU043360.D
Acq On : 27 Apr 2021 22:09
Operator : SY/MD
Sample : M2133-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ118I-20210420

Quant Time: Apr 28 09:12:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

