

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
 Data File : VU043363.D
 Acq On : 27 Apr 2021 23:18
 Operator : SY/MD
 Sample : M2133-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ125-20210420

Quant Time: Apr 28 09:13:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	102839	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	171273	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	163759	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	76587	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	92496	53.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.00%	
35) Dibromofluoromethane	5.298	113	63329	49.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.42%	
50) Toluene-d8	7.906	98	216404	47.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.04%	
62) 4-Bromofluorobenzene	10.639	95	75832	41.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.16%	
Target Compounds						
16) Acetone	2.642	43	1948	1.61	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
Data File : VU043363.D
Acq On : 27 Apr 2021 23:18
Operator : SY/MD
Sample : M2133-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ125-20210420

Quant Time: Apr 28 09:13:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

