

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
 Data File : VU043366.D
 Acq On : 28 Apr 2021 00:27
 Operator : SY/MD
 Sample : M2133-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ124-20210420

Quant Time: Apr 28 09:13:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	102943	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	174532	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	165111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	76901	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	91762	53.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.04%	
35) Dibromofluoromethane	5.298	113	64526	49.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.40%	
50) Toluene-d8	7.906	98	217872	46.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.90%	
62) 4-Bromofluorobenzene	10.639	95	75442	40.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.20%	
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.584	96	942	0.98	ug/l	88
16) Acetone	2.636	43	3056	2.70	ug/l	96
24) 1,1-Dichloroethane	3.880	63	18676	8.26	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
Data File : VU043366.D
Acq On : 28 Apr 2021 00:27
Operator : SY/MD
Sample : M2133-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ124-20210420

Quant Time: Apr 28 09:13:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

