

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
 Data File : VU043368.D
 Acq On : 28 Apr 2021 01:13
 Operator : SY/MD
 Sample : M2133-12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ147-20210420

Quant Time: Apr 28 09:14:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	105497	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	173981	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	167557	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	80108	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	93920	52.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.92%	
35) Dibromofluoromethane	5.298	113	65214	49.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.76%	
50) Toluene-d8	7.906	98	223033	48.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.42%	
62) 4-Bromofluorobenzene	10.639	95	78600	42.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.86%	
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.584	96	2845	2.88	ug/l	88
16) Acetone	2.632	43	2504	2.10	ug/l	100
24) 1,1-Dichloroethane	3.880	63	29274	12.63	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
Data File : VU043368.D
Acq On : 28 Apr 2021 01:13
Operator : SY/MD
Sample : M2133-12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ147-20210420

Quant Time: Apr 28 09:14:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

