

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
 Data File : VU043389.D
 Acq On : 28 Apr 2021 16:27
 Operator : SY/MD
 Sample : M2133-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ124-20210420

Quant Time: Apr 29 08:39:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	101485	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	169094	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	166624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	77919	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	97574	57.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.38%	
35) Dibromofluoromethane	5.298	113	67825	53.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.76%	
50) Toluene-d8	7.906	98	229678	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	10.639	95	81433	45.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.46%	
Target Compounds					Qvalue	
6) Chloroethane	1.938	64	1937	2.42	ug/l	96
12) 1,1-Dichloroethene	2.588	96	993	1.05	ug/l	91
16) Acetone	2.626	43	2535	2.22	ug/l	92
24) 1,1-Dichloroethane	3.880	63	18769	8.42	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
Data File : VU043389.D
Acq On : 28 Apr 2021 16:27
Operator : SY/MD
Sample : M2133-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ124-20210420

Quant Time: Apr 29 08:39:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

