

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
 Data File : VU043397.D
 Acq On : 28 Apr 2021 19:32
 Operator : SY/MD
 Sample : M2208-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW8-GW-358-360

Quant Time: Apr 29 08:42:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	108008	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	180241	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	174526	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	82158	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	101968	56.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.32%	
35) Dibromofluoromethane	5.295	113	71116	52.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.02%	
50) Toluene-d8	7.906	98	245270	51.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.36%	
62) 4-Bromofluorobenzene	10.639	95	84782	44.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.36%	
Target Compounds						
16) Acetone	2.630	43	4174	3.61	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
Data File : VU043397.D
Acq On : 28 Apr 2021 19:32
Operator : SY/MD
Sample : M2208-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW8-GW-358-360

Quant Time: Apr 29 08:42:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

