

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
 Data File : VU043401.D
 Acq On : 28 Apr 2021 21:05
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 29 08:43:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.516	0.508	1.6	80	0.00
3 P	Chloromethane	0.615	0.616	-0.2	90	0.00
4 C	Vinyl Chloride	0.563	0.559	0.7#	83	0.00
5 T	Bromomethane	0.307	0.350	-14.0	89	0.00
6 T	Chloroethane	0.432	0.358	17.1	75	0.00
7 T	Trichlorofluoromethane	0.956	0.891	6.8	78	0.00
8 T	Diethyl Ether	0.306	0.309	-1.0	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.492	0.511	-3.9	85	0.00
10 T	Methyl Iodide	0.648	0.649	-0.2	77	0.00
11 T	Tert butyl alcohol	0.217	0.212	2.3	80	-0.04
12 CM	1,1-Dichloroethene	0.468	0.491	-4.9#	86	0.00
13 T	Acrolein	0.125	0.139	-11.2	84	0.00
14 T	Allyl chloride	1.136	1.220	-7.4	91	0.00
15 T	Acrylonitrile	0.466	0.479	-2.8	86	0.00
16 T	Acetone	0.530	0.494	6.8	82	0.00
17 T	Carbon Disulfide	1.440	1.306	9.3	77	0.00
18 T	Methyl Acetate	1.051	1.264	-20.3#	101	0.00
19 T	Methyl tert-butyl Ether	1.849	2.026	-9.6	89	0.00
20 T	Methylene Chloride	0.597	0.630	-5.5	90	0.00
21 T	trans-1,2-Dichloroethene	0.547	0.544	0.5	83	0.00
22 T	Diisopropyl ether	2.173	2.434	-12.0	91	0.00
23 T	Vinyl Acetate	2.103	2.270	-7.9	86	0.00
24 P	1,1-Dichloroethane	1.099	1.192	-8.5	90	0.00
25 T	2-Butanone	0.764	0.778	-1.8	83	0.00
26 T	2,2-Dichloropropane	0.964	0.797	17.3	67	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.658	-5.8	85	0.00
28 T	Bromochloromethane	0.560	0.619	-10.5	95	0.00
29 T	Tetrahydrofuran	0.481	0.495	-2.9	84	0.00
30 C	Chloroform	1.164	1.243	-6.8#	87	0.00
31 T	Cyclohexane	1.037	0.987	4.8	80	0.00
32 T	1,1,1-Trichloroethane	1.029	1.061	-3.1	83	0.00
33 S	1,2-Dichloroethane-d4	0.841	0.885	-5.2	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.376	0.385	-2.4	90	0.00
36 T	1,1-Dichloropropene	0.530	0.519	2.1	81	0.00
37 T	Ethyl Acetate	0.907	0.895	1.3	86	0.00
38 T	Carbon Tetrachloride	0.569	0.568	0.2	82	0.00
39 T	Methylcyclohexane	0.583	0.528	9.4	74	0.00
40 TM	Benzene	1.516	1.571	-3.6	86	0.00
41 T	Methacrylonitrile	0.477	0.488	-2.3	88	0.00
42 TM	1,2-Dichloroethane	0.672	0.702	-4.5	87	0.00
43 T	Isopropyl Acetate	1.294	1.337	-3.3	86	0.00
44 TM	Trichloroethene	0.384	0.402	-4.7	87	0.00
45 C	1,2-Dichloropropane	0.422	0.439	-4.0#	86	0.00
46 T	Dibromomethane	0.299	0.307	-2.7	86	0.00
47 T	Bromodichloromethane	0.590	0.628	-6.4	88	0.00
48 T	Methyl methacrylate	0.649	0.678	-4.5	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
 Data File : VU043401.D
 Acq On : 28 Apr 2021 21:05
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 29 08:43:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.013	0.013	0.0	75	0.00
50 S	Toluene-d8	1.329	1.314	1.1	85	0.00
51 T	4-Methyl-2-Pentanone	0.957	0.953	0.4	82	0.00
52 CM	Toluene	0.940	0.961	-2.2#	83	0.00
53 T	t-1,3-Dichloropropene	0.655	0.624	4.7	77	0.00
54 T	cis-1,3-Dichloropropene	0.665	0.658	1.1	81	0.00
55 T	1,1,2-Trichloroethane	0.403	0.421	-4.5	87	0.00
56 T	Ethyl methacrylate	0.660	0.640	3.0	78	0.00
57 T	1,3-Dichloropropane	0.696	0.725	-4.2	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.235	0.224	4.7	72	0.00
59 T	2-Hexanone	0.788	0.764	3.0	78	0.00
60 T	Dibromochloromethane	0.466	0.480	-3.0	83	0.00
61 T	1,2-Dibromoethane	0.455	0.452	0.7	81	0.00
62 S	4-Bromofluorobenzene	0.532	0.525	1.3	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.363	0.401	-10.5	94	0.00
65 PM	Chlorobenzene	1.072	1.024	4.5	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.410	0.412	-0.5	82	0.00
67 C	Ethyl Benzene	1.935	1.926	0.5#	82	0.00
68 T	m/p-Xylenes	0.716	0.711	0.7	80	0.00
69 T	o-Xylene	0.699	0.684	2.1	80	0.00
70 T	Styrene	1.177	1.188	-0.9	80	0.00
71 P	Bromoform	0.422	0.395	6.4	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
73 T	Isopropylbenzene	3.356	3.339	0.5	79	0.00
74 T	N-amyl acetate	2.156	1.972	8.5	73	0.00
75 P	1,1,2,2-Tetrachloroethane	1.461	1.416	3.1	77	0.00
76 T	1,2,3-Trichloropropane	1.461	1.337	8.5	77	0.00
77 T	Bromobenzene	0.935	0.890	4.8	79	0.00
78 T	n-propylbenzene	4.018	3.912	2.6	76	0.00
79 T	2-Chlorotoluene	2.442	2.398	1.8	78	0.00
80 T	1,3,5-Trimethylbenzene	2.962	2.896	2.2	74	0.00
81 T	trans-1,4-Dichloro-2-butene	0.497	0.397	20.1#	62	0.00
82 T	4-Chlorotoluene	2.933	2.859	2.5	78	0.00
83 T	tert-Butylbenzene	2.912	2.729	6.3	73	0.00
84 T	1,2,4-Trimethylbenzene	3.044	2.931	3.7	74	0.00
85 T	sec-Butylbenzene	3.413	3.217	5.7	73	0.00
86 T	p-Isopropyltoluene	3.142	2.935	6.6	72	0.00
87 T	1,3-Dichlorobenzene	1.695	1.611	5.0	76	0.00
88 T	1,4-Dichlorobenzene	1.768	1.609	9.0	73	0.00
89 T	n-Butylbenzene	2.943	2.603	11.6	67	0.00
90 T	Hexachloroethane	0.589	0.457	22.4#	62	0.00
91 T	1,2-Dichlorobenzene	1.732	1.646	5.0	76	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.414	0.369	10.9	70	0.00
93 T	1,2,4-Trichlorobenzene	1.156	1.075	7.0	78	0.00
94 T	Hexachlorobutadiene	0.624	0.502	19.6	68	0.00
95 T	Naphthalene	3.495	3.634	-4.0	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042821\
Data File : VU043401.D
Acq On : 28 Apr 2021 21:05
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Apr 29 08:43:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.135	1.100	3.1	81	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6