

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042822\
 Data File : VU048331.D
 Acq On : 28 Apr 2022 11:19
 Operator : SY/MD
 Sample : VU0428WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK106

Quant Time: Apr 29 04:28:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 29 04:27:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	351636	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	352168	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	170550	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	141953	43.900	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.800%
7) Chloroethane-d5	1.903	69	124401	54.808	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.620%
11) 1,1-Dichloroethene-d2	2.565	63	193952	37.435	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.860%
21) 2-Butanone-d5	4.626	46	273734	99.254	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.250%
24) Chloroform-d	5.063	84	284508	48.509	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.020%
26) 1,2-Dichloroethane-d4	5.703	65	181789	52.239	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.480%
32) Benzene-d6	5.729	84	599467	53.972	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.940%
36) 1,2-Dichloropropane-d6	6.690	67	193247	52.628	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.260%
41) Toluene-d8	7.899	98	535338	53.174	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.340%
43) trans-1,3-Dichloroprop...	8.179	79	83437	51.495	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.000%
47) 2-Hexanone-d5	8.636	63	202886	102.410	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.410%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	310105	48.153	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.300%
66) 1,2-Dichlorobenzene-d4	12.195	152	214076	55.799	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042822\
Data File : VU048331.D
Acq On : 28 Apr 2022 11:19
Operator : SY/MD
Sample : VU0428WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK106

Quant Time: Apr 29 04:28:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 29 04:27:14 2022
Response via : Initial Calibration

