

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042919\
 Data File : VU031598.D
 Acq On : 29 Apr 2019 19:47
 Operator : JC/SP
 Sample : K2536-22
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0X10

Manual Integrations
 APPROVED

MMDadoda
 4/30/2019 12:07:42 PM

Quant Time: Apr 30 04:39:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 30 01:54:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	658670	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	646406	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	239405	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	285360	51.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.54%
7) Chloroethane-d5	1.68	69	237508	55.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	111.50%
11) 1,1-Dichloroethene-d2	2.27	63	441018	40.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.34%
21) 2-Butanone-d5	4.18	46	596727	119.47	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.47%
24) Chloroform-d	4.64	84	532433	52.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.48%
26) 1,2-Dichloroethane-d4	5.31	65	369791	56.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.12%
32) Benzene-d6	5.34	84	1036956	52.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.48%
36) 1,2-Dichloropropane-d6	6.32	67	379426	54.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.66%
41) Toluene-d8	7.56	98	857457	49.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.20%
43) trans-1,3-Dichloropropene-	7.85	79	138750	46.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.56%
47) 2-Hexanone-d5	8.31	63	332486	110.00	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.00%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	478717	51.08	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.16%
64) 1,2-Dichlorobenzene-d4	11.86	152	273252	54.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.54%

Target Compounds

					Ovalue
9) Trichlorofluoromethane	1.88	101	204769	21.924 ug/L	98
13) Acetone	2.34	43	86548	15.383 ug/L	95
20) cis-1,2-Dichloroethene	4.23	96	6420m	1.118 ug/L	
31) Carbon tetrachloride	5.13	117	4460	0.576 ug/L	90
34) Trichloroethene	6.18	95	3366724	584.410 ug/L	95
46) Tetrachloroethene	8.22	164	310085	78.954 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042919\
Data File : VU031598.D
Acq On : 29 Apr 2019 19:47
Operator : JC/SP
Sample : K2536-22
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 29 Sample Multiplier: 1


Instrument :
MSVOA_U
ClientSampleId :
C0X10

Manual Integrations
APPROVED

MMDadoda
4/30/2019 12:07:42 PM

Quant Time: Apr 30 04:39:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042719WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 30 01:54:25 2019
Response via : Initial Calibration









#9
 Trichlorofluoromethane
 Concen: 21.924 ug/L
 RT: 1.88 min Scan# 247
 Delta R.T. 0.00 min
 Lab File: VU031598.D
 Acq: 29 Apr 2019 19:47

Instrument :
 MSVOA_U
 ClientSampleId :
 C0X10

Tot Ion	Ratio	Resp	Lower	Upper
101	100	204769		
103	65.4	50.8	76.2	




Manual Integrations
 APPROVED


MMDadoda
 4/30/2019 12:07:42 PM



#13
 Acetone
 Concen: 15.383 ug/L
 RT: 2.34 min Scan# 388
 Delta R.T. 0.00 min
 Lab File: VU031598.D
 Acq: 29 Apr 2019 19:47

Tgt Ion	Ratio	Resp	Lower	Upper
43	100	86548		
58	29.7	0.0	64.4	





#20


cis-1,2-Dichloroethene
Concen: 1.118 ug/L m
RT: 4.23 min Scan# 975
Delta R.T. 0.01 min
Lab File: VU031598.D
Acq: 29 Apr 2019 19:47

Instrument :
MSVOA_U
ClientSampleId :
C0X10




Tot Ion	Ratio	Lower	Upper
96	100		
61	169.6	99.5	184.7
68	0.0	0.0	0.0

Manual Integrations
APPROVED

MMDadoda
4/30/2019 12:07:42 PM




Abundance Ion 96.00 (95.70 to 96.70): VU031571.D (-956) (-)





#31


Carbon tetrachloride
Concen: 0.576 ug/L
RT: 5.13 min Scan# 1256
Delta R.T. 0.00 min
Lab File: VU031598.D
Acq: 29 Apr 2019 19:47

Tgt Ion	Ratio	Lower	Upper
117	100		
119	105.3	76.6	114.8



Abundance Ion 117.00 (116.70 to 117.70): VU031571.D (-1237) (-)





#34

Trichloroethene

Concen: 584.410 ug/L

RT: 6.18 min Scan# 1583

Delta R.T. 0.00 min

Lab File: VU031598.D

Acq: 29 Apr 2019 19:47

Instrument :

MSVOA_U

ClientSampled :

C0X10

Tot Ion: 95 Resp: 3366724

Ion Ratio Lower Upper

95 100

97 63.7 44.8 83.2

132 83.8 63.1 117.1





130 88.1 66.5 123.5

Manual Integrations

APPROVED

MMDadoda

4/30/2019 12:07:42 PM



#46

Tetrachloroethene

Concen: 78.954 ug/L

RT: 8.22 min Scan# 2218

Delta R.T. 0.00 min

Lab File: VU031598.D

Acq: 29 Apr 2019 19:47

Tgt Ion: 164 Resp: 310085

Ion Ratio Lower Upper

164 100

129 100.1 71.5 132.9

131 95.7 69.0 128.2

166 124.5 93.7 173.9

