

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042920\
 Data File : VU037917.D
 Acq On : 29 Apr 2020 17:21
 Operator : JC/MD
 Sample : L2410-15
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBFF3

Quant Time: Apr 30 03:19:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 30 01:17:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	520494	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	510973	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	231435	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	149572	32.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.78%
7) Chloroethane-d5	1.93	69	144457	40.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.96%
11) 1,1-Dichloroethene-d2	2.59	63	204789	27.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.64%#
21) 2-Butanone-d5	4.66	46	402138	111.60	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.60%
24) Chloroform-d	5.10	84	312785	45.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.30%
26) 1,2-Dichloroethane-d4	5.74	65	219947	45.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.72%
32) Benzene-d6	5.76	84	639496	44.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.52%
36) 1,2-Dichloropropane-d6	6.72	67	223736	46.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.58%
41) Toluene-d8	7.92	98	571636	43.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.08%
43) trans-1,3-Dichloropropene-	8.20	79	104507	42.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.02%
47) 2-Hexanone-d5	8.65	63	304612	115.78	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	115.78%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	344263	51.20	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.40%
64) 1,2-Dichlorobenzene-d4	12.20	152	215543	47.86	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.72%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.60	96	2725	0.862 ug/L #	1
13) Acetone	2.65	43	9494	4.416 ug/L	97
15) Methyl Acetate	2.98	43	5187	0.998 ug/L #	73
20) cis-1,2-Dichloroethene	4.71	96	382147	98.783 ug/L	94
34) Trichloroethene	6.57	95	208438	54.791 ug/L	98
46) Tetrachloroethene	8.57	164	451266	162.293 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042920\
Data File : VU037917.D
Acq On : 29 Apr 2020 17:21
Operator : JC/MD
Sample : L2410-15
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBFF3

Quant Time: Apr 30 03:19:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042320WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 30 01:17:58 2020
Response via : Initial Calibration

