

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU042920\
 Data File : VU037929.D
 Acq On : 29 Apr 2020 22:40
 Operator : JC/MD
 Sample : L2410-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBEF2

Quant Time: Apr 30 05:51:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 30 04:10:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	626243	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	594269	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	262102	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	245105	44.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.58%
7) Chloroethane-d5	1.93	69	211392	49.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.70%
11) 1,1-Dichloroethene-d2	2.59	63	303288	34.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.48%
21) 2-Butanone-d5	4.66	46	410790	94.75	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.75%
24) Chloroform-d	5.10	84	398774	47.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.70%
26) 1,2-Dichloroethane-d4	5.74	65	292763	50.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.46%
32) Benzene-d6	5.76	84	870545	51.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.60%
36) 1,2-Dichloropropane-d6	6.72	67	292741	52.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.16%
41) Toluene-d8	7.92	98	753180	48.76	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.52%
43) trans-1,3-Dichloropropene-	8.20	79	109582	37.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.76%
47) 2-Hexanone-d5	8.65	63	277314	90.63	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.63%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	405522	51.85	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.70%
64) 1,2-Dichlorobenzene-d4	12.20	152	264352	51.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.66%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.59	96	3682	0.968 ug/L #	1
13) Acetone	2.65	43	8731	3.375 ug/L	93
20) cis-1,2-Dichloroethene	4.71	96	117781	25.305 ug/L	92
34) Trichloroethene	6.57	95	394247	89.109 ug/L	98
46) Tetrachloroethene	8.57	164	2527267	781.508 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU042920\
Data File : VU037929.D
Acq On : 29 Apr 2020 22:40
Operator : JC/MD
Sample : L2410-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBEF2

Quant Time: Apr 30 05:51:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM042320WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 30 04:10:11 2020
Response via : Initial Calibration

