

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042921\
 Data File : VU043403.D
 Acq On : 28 Apr 2021 22:14
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Apr 29 09:44:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	86797	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	140854	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	134686	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	77040	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	76584	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
35) Dibromofluoromethane	5.295	113	53790	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
50) Toluene-d8	7.906	98	184819	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
62) 4-Bromofluorobenzene	10.639	95	72620	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.385	85	45595	50.91	ug/l	98
3) Chloromethane	1.523	50	54422	57.20	ug/l	98
4) Vinyl Chloride	1.607	62	48337	49.50	ug/l	97
5) Bromomethane	1.835	94	30901	58.07	ug/l	95
6) Chloroethane	1.922	64	31940	47.41	ug/l	97
7) Trichlorofluoromethane	2.134	101	78904	47.55	ug/l	100
8) Diethyl Ether	2.379	74	27993	52.68	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.581	101	44612	52.25	ug/l	97
10) Methyl Iodide	2.726	142	57248	50.88	ug/l	98
11) Tert butyl alcohol	3.189	59	92449	245.83	ug/l #	77
12) 1,1-Dichloroethene	2.581	96	42703	52.55	ug/l	95
13) Acrolein	2.488	56	62686	266.68	ug/l	98
14) Allyl chloride	2.925	41	105625	53.57	ug/l	96
15) Acrylonitrile	3.317	53	208893	258.23	ug/l	99
16) Acetone	2.629	43	214666	249.98	ug/l	99
17) Carbon Disulfide	2.793	76	116251	46.52	ug/l	97
18) Methyl Acetate	2.951	43	110904	60.81	ug/l	96
19) Methyl tert-butyl Ether	3.369	73	177675	55.34	ug/l	99
20) Methylene Chloride	3.051	84	56075	54.11	ug/l	99
21) trans-1,2-Dichloroethene	3.356	96	48387	50.98	ug/l	97
22) Diisopropyl ether	3.999	45	214726	56.93	ug/l	90
23) Vinyl Acetate	3.960	43	991560	271.64	ug/l	99
24) 1,1-Dichloroethane	3.874	63	105939	55.54	ug/l	99
25) 2-Butanone	4.706	43	336720	254.03	ug/l	100
26) 2,2-Dichloropropane	4.671	77	70268	42.00	ug/l	99
27) cis-1,2-Dichloroethene	4.674	96	57179	52.94	ug/l	97
28) Bromochloromethane	4.980	49	53942	55.46	ug/l	91
29) Tetrahydrofuran	5.054	42	215393	258.07	ug/l	96
30) Chloroform	5.095	83	109845	54.38	ug/l	94
31) Cyclohexane	5.394	56	87434	48.58	ug/l	92
32) 1,1,1-Trichloroethane	5.324	97	93408	52.31	ug/l	98
36) 1,1-Dichloropropene	5.533	75	72305	48.39	ug/l	100
37) Ethyl Acetate	4.809	43	121591	47.58	ug/l	98
38) Carbon Tetrachloride	5.530	117	79034	49.33	ug/l	99
39) Methylcyclohexane	6.771	83	72513	44.14	ug/l	95
40) Benzene	5.780	78	217608	50.96	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042921\
 Data File : VU043403.D
 Acq On : 28 Apr 2021 22:14
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Apr 29 09:44:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.980	41	68307	50.85	ug/l	98
42) 1,2-Dichloroethane	5.800	62	99182	52.37	ug/l	99
43) Isopropyl Acetate	5.919	43	183788	50.44	ug/l	98
44) Trichloroethene	6.549	130	56052	51.76	ug/l	90
45) 1,2-Dichloropropane	6.796	63	63681	53.53	ug/l	99
46) Dibromomethane	6.922	93	42604	50.64	ug/l	96
47) Bromodichloromethane	7.111	83	87375	52.61	ug/l	99
48) Methyl methacrylate	6.967	41	94240	51.52	ug/l	96
49) 1,4-Dioxane	6.967	88	34337	936.80	ug/l #	95
51) 4-Methyl-2-Pentanone	7.796	43	662593	245.77	ug/l	97
52) Toluene	7.976	92	135513	51.19	ug/l	99
53) t-1,3-Dichloropropene	8.218	75	87150	47.26	ug/l	99
54) cis-1,3-Dichloropropene	7.613	75	92382	49.34	ug/l	96
55) 1,1,2-Trichloroethane	8.407	97	59170	52.11	ug/l	99
56) Ethyl methacrylate	8.340	69	88579	47.63	ug/l	95
57) 1,3-Dichloropropane	8.581	76	102310	52.16	ug/l	100
58) 2-Chloroethyl Vinyl ether	7.468	63	155042	201.99	ug/l	98
59) 2-Hexanone	8.690	43	525775	236.84	ug/l	98
60) Dibromochloromethane	8.816	129	67232	51.19	ug/l	100
61) 1,2-Dibromoethane	8.931	107	64492	50.35	ug/l	98
64) Tetrachloroethene	8.558	164	56866	58.19	ug/l	98
65) Chlorobenzene	9.452	112	143047	49.53	ug/l	99
66) 1,1,1,2-Tetrachloroethane	9.539	131	57286	51.88	ug/l	99
67) Ethyl Benzene	9.578	91	264066	50.66	ug/l	99
68) m/p-Xylenes	9.700	106	194647	100.92	ug/l	99
69) o-Xylene	10.108	106	93488	49.66	ug/l	99
70) Styrene	10.121	104	163573	51.61	ug/l	99
71) Bromoform	10.298	173	55444	48.73	ug/l #	100
73) Isopropylbenzene	10.491	105	255449	49.40	ug/l	99
74) N-amyl acetate	10.327	43	151719	45.67	ug/l	97
75) 1,1,2,2-Tetrachloroethane	10.790	83	110127	48.91	ug/l	99
76) 1,2,3-Trichloropropane	10.832	75	122201	54.28	ug/l	83
77) Bromobenzene	10.790	156	69642	48.35	ug/l	99
78) n-propylbenzene	10.912	91	309117	49.93	ug/l	99
79) 2-Chlorotoluene	10.992	91	184002	48.91	ug/l	100
80) 1,3,5-Trimethylbenzene	11.095	105	225398	49.38	ug/l	99
81) trans-1,4-Dichloro-2-b...	10.552	75	29309	38.26	ug/l	90
82) 4-Chlorotoluene	11.102	91	221176	48.94	ug/l	99
83) tert-Butylbenzene	11.426	119	213032	47.47	ug/l	98
84) 1,2,4-Trimethylbenzene	11.475	105	230263	49.10	ug/l	99
85) sec-Butylbenzene	11.651	105	251933	47.91	ug/l	98
86) p-Isopropyltoluene	11.799	119	231071	47.73	ug/l	99
87) 1,3-Dichlorobenzene	11.751	146	122706	46.99	ug/l	98
88) 1,4-Dichlorobenzene	11.841	146	126643	46.50	ug/l	99
89) n-Butylbenzene	12.214	91	206958	45.64	ug/l	98
90) Hexachloroethane	12.484	117	35837	39.49	ug/l	93
91) 1,2-Dichlorobenzene	12.217	146	128290	48.08	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	13.002	75	27708	43.47	ug/l	92
93) 1,2,4-Trichlorobenzene	13.848	180	82466	46.84	ug/l	97
94) Hexachlorobutadiene	14.028	225	37932	42.89	ug/l	99
95) Naphthalene	14.095	128	277580	49.82	ug/l	100
96) 1,2,3-Trichlorobenzene	14.336	180	86088	50.32	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042921\
Data File : VU043403.D
Acq On : 28 Apr 2021 22:14
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Apr 29 09:44:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042921\
 Data File : VU043403.D
 Acq On : 28 Apr 2021 22:14
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Apr 29 09:44:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.382 min Scan# 1257
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:168 Resp: 86797
Ion Ratio Lower Upper
168 100
99 59.2 45.0 67.4



#2
Dichlorodifluoromethane
Concen: 50.91 ug/l
RT: 1.385 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 85 Resp: 45595
Ion Ratio Lower Upper
85 100
87 31.8 16.5 49.5





#3

Chloromethane

Concen: 57.20 ug/l

RT: 1.523 min Scan# 57





Delta R.T. 0.003 min

Lab File: VU043403.D

Acq: 28 Apr 2021 22:14

Tgt Ion: 50 Resp: 54422

Ion	Ratio	Lower	Upper
50	100		
52	32.1	26.6	40.0



#4

Vinyl Chloride

Concen: 49.50 ug/l

RT: 1.607 min Scan# 83




Delta R.T. -0.000 min

Lab File: VU043403.D

Acq: 28 Apr 2021 22:14

Tgt Ion: 62 Resp: 48337

Ion	Ratio	Lower	Upper
62	100		
64	33.0	25.2	37.8









#5
Bromomethane
Concen: 58.07 ug/l
RT: 1.835 min Scan# 154
Delta R.T. -0.007 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 94 Resp: 30901
Ion Ratio Lower Upper
94 100
96 90.6 76.6 115.0



#6
Chloroethane
Concen: 47.41 ug/l
RT: 1.922 min Scan# 181
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 64 Resp: 31940
Ion Ratio Lower Upper
64 100
66 31.6 26.4 39.6









#7
Trichlorofluoromethane
Concen: 47.55 ug/l
RT: 2.134 min Scan# 247
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :

Tgt Ion:101 Resp: 78904
Ion Ratio Lower Upper
101 100
103 64.1 51.3 76.9



#8
Diethyl Ether
Concen: 52.68 ug/l
RT: 2.379 min Scan# 323
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 74 Resp: 27993
Ion Ratio Lower Upper
74 100
45 119.9 56.9 170.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.25 ug/l
RT: 2.581 min Scan# 386
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:101 Resp: 44612
Ion Ratio Lower Upper
101 100
85 47.9 37.4 56.0
151 79.7 65.8 98.8



#10
Methyl Iodide
Concen: 50.88 ug/l
RT: 2.726 min Scan# 431
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion:142 Resp: 57248
Ion Ratio Lower Upper
142 100
127 48.9 37.8 56.6
141 15.0 11.8 17.6









#11
Tert butyl alcohol
Concen: 245.83 ug/l
RT: 3.189 min Scan# 575
Delta R.T. -0.035 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 59 Resp: 92449
Ion Ratio Lower Upper
59 100
57 10.3 1.6 2.4#



#12
1,1-Dichloroethene
Concen: 52.55 ug/l
RT: 2.581 min Scan# 386
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14


Tgt Ion: 96 Resp: 42703
Ion Ratio Lower Upper
96 100
61 193.7 147.4 221.0
98 65.5 51.4 77.2







#13
Acrolein
Concen: 266.68 ug/l
RT: 2.488 min Scan# 357
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14


Instrument :
MSVOA_U
ClientSampleId :





Tgt Ion: 56 Resp: 62686
Ion Ratio Lower Upper
56 100
55 73.2 57.3 85.9




#14
Allyl chloride
Concen: 53.57 ug/l
RT: 2.925 min Scan# 493
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Tgt Ion: 41 Resp: 105625
Ion Ratio Lower Upper
41 100
39 66.4 54.9 82.3
76 25.2 22.5 33.7








#15
Acrylonitrile
Concen: 258.23 ug/l
RT: 3.317 min Scan# 615
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14


Instrument :
MSVOA_U
ClientSampled :





Tgt Ion: 53 Resp: 208893
Ion Ratio Lower Upper
53 100
52 85.2 68.2 102.2
51 38.1 29.2 43.8




#16
Acetone
Concen: 249.98 ug/l
RT: 2.629 min Scan# 401
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14



Tgt Ion: 43 Resp: 214666
Ion Ratio Lower Upper
43 100
58 26.3 21.5 32.3









#17
Carbon Disulfide
Concen: 46.52 ug/l
RT: 2.793 min Scan# 452
Delta R.T. -0.004 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 76 Resp: 116251
Ion Ratio Lower Upper
76 100
78 8.4 7.5 11.3



#18
Methyl Acetate
Concen: 60.81 ug/l
RT: 2.951 min Scan# 501
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 43 Resp: 110904
Ion Ratio Lower Upper
43 100
74 18.9 16.6 25.0









#19

Methyl tert-butyl Ether
Concen: 55.34 ug/l
RT: 3.369 min Scan# 631
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Instrument :
MSVOA_U
ClientSampled :




Tgt Ion: 73 Resp: 177675
Ion Ratio Lower Upper
73 100
57 23.2 18.8 28.2




#20

Methylene Chloride
Concen: 54.11 ug/l
RT: 3.051 min Scan# 532
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 84 Resp: 56075
Ion Ratio Lower Upper
84 100
49 142.4 112.6 169.0
51 43.1 33.7 50.5
86 63.4 51.7 77.5









#21
trans-1,2-Dichloroethene
Concen: 50.98 ug/l
RT: 3.356 min Scan# 627
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 96 Resp: 48387
Ion Ratio Lower Upper
96 100
61 157.2 122.4 183.6
98 62.1 51.9 77.9



#22
Diisopropyl ether
Concen: 56.93 ug/l
RT: 3.999 min Scan# 827
Delta R.T. 0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 45 Resp: 214726
Ion Ratio Lower Upper
45 100
43 79.9 73.0 109.6
87 22.3 19.5 29.3
59 10.4 8.6 12.8









#23
 Vinyl Acetate
 Concen: 271.64 ug/l
 RT: 3.960 min Scan# 815
 Delta R.T. -0.001 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14




Instrument :
 MSVOA_U
 ClientSampled :


Tgt Ion: 43 Resp: 991560
 Ion Ratio Lower Upper
 43 100
 86 8.0 6.7 10.1



#24
 1,1-Dichloroethane
 Concen: 55.54 ug/l
 RT: 3.874 min Scan# 788
 Delta R.T. -0.003 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14

Tgt Ion: 63 Resp: 105939
 Ion Ratio Lower Upper
 63 100
 98 6.1 3.1 9.3
 100 3.5 2.1 6.2









#25
2-Butanone
Concen: 254.03 ug/l
RT: 4.706 min Scan# 1047
Delta R.T. -0.001 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :

Tgt Ion: 43 Resp: 336720
Ion Ratio Lower Upper
43 100
72 21.6 17.4 26.0



#26
2,2-Dichloropropane
Concen: 42.00 ug/l
RT: 4.671 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 77 Resp: 70268
Ion Ratio Lower Upper
77 100
97 21.7 10.6 31.8









#27
 cis-1,2-Dichloroethene
 Concen: 52.94 ug/l
 RT: 4.674 min Scan# 1037
 Delta R.T. 0.000 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14




Instrument :
 MSVOA_U
 ClientSampled :


Tgt Ion:	96	Resp:	57179
Ion	Ratio	Lower	Upper
96	100		
61	165.3	0.0	322.2
98	65.0	0.0	128.6



#28
 Bromochloromethane
 Concen: 55.46 ug/l
 RT: 4.980 min Scan# 1132
 Delta R.T. -0.000 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14

Tgt Ion:	49	Resp:	53942
Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	3.4
130	64.0	57.1	85.7





#29

Tetrahydrofuran

Concen: 258.07 ug/l

RT: 5.054 min Scan# 1155





Delta R.T. -0.000 min

Lab File: VU043403.D

Acq: 28 Apr 2021 22:14

Tgt Ion: 42 Resp: 215393

Ion	Ratio	Lower	Upper
42	100		
72	36.1	31.5	47.3
71	34.9	29.5	44.3



#30

Chloroform

Concen: 54.38 ug/l

RT: 5.095 min Scan# 1168




Delta R.T. -0.001 min


Lab File: VU043403.D

Acq: 28 Apr 2021 22:14

Tgt Ion: 83 Resp: 109845

Ion	Ratio	Lower	Upper
83	100		
85	63.7	54.6	81.8









#31
Cyclohexane
Concen: 48.58 ug/l
RT: 5.394 min Scan# 1261
Delta R.T. -0.001 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 56 Resp: 87434
Ion Ratio Lower Upper
56 100
69 28.6 24.9 37.3
84 71.9 63.7 95.5



#32
1,1,1-Trichloroethane
Concen: 52.31 ug/l
RT: 5.324 min Scan# 1239
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 97 Resp: 93408
Ion Ratio Lower Upper
97 100
99 64.9 51.4 77.0
61 49.0 37.4 56.2









#33
1,2-Dichloroethane-d4
Concen: 52.49 ug/l
RT: 5.710 min Scan# 1359
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 65 Resp: 76584
Ion Ratio Lower Upper
65 100
67 50.5 0.0 99.8



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.256 min Scan# 1529
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 114 Resp: 140854
Ion Ratio Lower Upper
114 100
63 23.9 0.0 44.2
88 16.5 0.0 33.4









#35
Dibromofluoromethane
Concen: 50.82 ug/l
RT: 5.295 min Scan# 1230
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:113 Resp: 53790
Ion Ratio Lower Upper
113 100
111 101.3 81.8 122.8
192 20.1 17.2 25.8



#36
1,1-Dichloropropene
Concen: 48.39 ug/l
RT: 5.533 min Scan# 1304
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 75 Resp: 72305
Ion Ratio Lower Upper
75 100
110 33.6 16.8 50.3
77 30.6 24.8 37.2





#37

Ethyl Acetate

Concen: 47.58 ug/l

RT: 4.809 min Scan# 1079





Delta R.T. 0.003 min

Lab File: VU043403.D

Acq: 28 Apr 2021 22:14

Tgt Ion: 43 Resp: 121591

Ion	Ratio	Lower	Upper
43	100		
61	12.3	10.1	15.1
70	8.1	7.2	10.8



#38

Carbon Tetrachloride

Concen: 49.33 ug/l

RT: 5.530 min Scan# 1303




Delta R.T. -0.003 min


Lab File: VU043403.D

Acq: 28 Apr 2021 22:14

Tgt Ion: 117 Resp: 79034

Ion	Ratio	Lower	Upper
117	100		
119	95.9	75.5	113.3
121	31.0	24.7	37.1





#39

Methylcyclohexane

Concen: 44.14 ug/l

RT: 6.771 min Scan# 1689

Delta R.T. -0.000 min

Lab File: VU043403.D

Acq: 28 Apr 2021 22:14





Tgt Ion: 83 Resp: 72513

Ion Ratio Lower Upper

83 100

55 91.1 66.8 100.2

98 46.0 36.9 55.3



#40

Benzene

Concen: 50.96 ug/l

RT: 5.780 min Scan# 1381

Delta R.T. 0.000 min

Lab File: VU043403.D




Acq: 28 Apr 2021 22:14

Tgt Ion: 78 Resp: 217608

Ion Ratio Lower Upper

78 100

77 24.6 19.3 28.9









#41
Methacrylonitrile
Concen: 50.85 ug/l
RT: 4.980 min Scan# 1132
Delta R.T. 0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 41 Resp: 68307
Ion Ratio Lower Upper
41 100
39 56.2 45.1 67.7
67 59.2 50.2 75.2
52 27.2 22.1 33.1



#42
1,2-Dichloroethane
Concen: 52.37 ug/l
RT: 5.800 min Scan# 1387
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 62 Resp: 99182
Ion Ratio Lower Upper
62 100
98 7.1 0.0 14.8









#43

Isopropyl Acetate
 Concen: 50.44 ug/l
 RT: 5.919 min Scan# 1424
 Delta R.T. 0.004 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14

Instrument :
 MSVOA_U
 ClientSampled :






Tgt Ion:	43	Resp:	183788
Ion	Ratio	Lower	Upper
43	100		
61	16.8	14.4	21.6
87	10.5	8.8	13.2




#44

Trichloroethene
 Concen: 51.76 ug/l
 RT: 6.549 min Scan# 1620
 Delta R.T. -0.000 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14

Tgt Ion:	130	Resp:	56052
Ion	Ratio	Lower	Upper
130	100		
95	108.7	0.0	197.8





#45

1,2-Dichloropropane

Concen: 53.53 ug/l

RT: 6.796 min Scan# 1697

Delta R.T. 0.000 min

Lab File: VU043403.D





Acq: 28 Apr 2021 22:14

Tgt Ion: 63 Resp: 63681

Ion Ratio Lower Upper

63 100

65 32.2 25.4 38.2



#46

Dibromomethane

Concen: 50.64 ug/l

RT: 6.922 min Scan# 1736

Delta R.T. -0.003 min

Lab File: VU043403.D

Acq: 28 Apr 2021 22:14




Tgt Ion: 93 Resp: 42604


Ion Ratio Lower Upper

93 100

95 84.7 63.6 95.4

174 101.1 83.4 125.2









#47
Bromodichloromethane
Concen: 52.61 ug/l
RT: 7.111 min Scan# 1795
Delta R.T. -0.001 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 83 Resp: 87375
Ion Ratio Lower Upper
83 100
85 64.5 51.3 76.9
127 8.0 6.9 10.3



#48
Methyl methacrylate
Concen: 51.52 ug/l
RT: 6.967 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14


Tgt Ion: 41 Resp: 94240
Ion Ratio Lower Upper
41 100
69 67.3 57.0 85.6
39 58.7 48.4 72.6








#49
1,4-Dioxane
Concen: 936.80 ug/l
RT: 6.967 min Scan# 1750
Delta R.T. -0.010 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14


Instrument :
MSVOA_U
ClientSampled :





Tgt Ion: 88 Resp: 34337
Ion Ratio Lower Upper
88 100
43 41.2 0.0 0.0#
58 79.9 60.4 90.6




#50
Toluene-d8
Concen: 49.35 ug/l
RT: 7.906 min Scan# 2042
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14



Tgt Ion: 98 Resp: 184819
Ion Ratio Lower Upper
98 100
100 62.9 50.2 75.2









#51
4-Methyl-2-Pentanone
Concen: 245.77 ug/l
RT: 7.796 min Scan# 2008
Delta R.T. -0.004 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 43 Resp: 662593
Ion Ratio Lower Upper
43 100
58 33.6 28.1 42.1



#52
Toluene
Concen: 51.19 ug/l
RT: 7.976 min Scan# 2064
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 92 Resp: 135513
Ion Ratio Lower Upper
92 100
91 173.9 140.3 210.5









#53
 t-1,3-Dichloropropene
 Concen: 47.26 ug/l
 RT: 8.218 min Scan# 2139
 Delta R.T. -0.000 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14




Instrument :
 MSVOA_U
 ClientSampled :


Tgt Ion: 75 Resp: 87150
 Ion Ratio Lower Upper
 75 100
 77 32.8 25.6 38.4



#54
 cis-1,3-Dichloropropene
 Concen: 49.34 ug/l
 RT: 7.613 min Scan# 1951
 Delta R.T. 0.000 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14


Tgt Ion: 75 Resp: 92382
 Ion Ratio Lower Upper
 75 100
 77 32.2 24.6 37.0
 39 61.5 46.1 69.1







#55
1,1,2-Trichloroethane
Concen: 52.11 ug/l
RT: 8.407 min Scan# 2198
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Instrument :
MSVOA_U
ClientSampled :







Tgt Ion:	97	Resp:	59170
Ion	Ratio	Lower	Upper
97	100		
83	89.6	72.6	109.0
85	58.4	46.6	69.8
99	62.9	50.9	76.3



#56
Ethyl methacrylate
Concen: 47.63 ug/l
RT: 8.340 min Scan# 2177
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion:	69	Resp:	88579
Ion	Ratio	Lower	Upper
69	100		
41	83.5	62.8	94.2
39	47.7	36.2	54.2









#57

1,3-Dichloropropane
Concen: 52.16 ug/l
RT: 8.581 min Scan# 2252
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Instrument :
MSVOA_U
ClientSampled :




Tgt Ion: 76 Resp: 102310
Ion Ratio Lower Upper
76 100
78 31.9 25.7 38.5




#58

2-Chloroethyl Vinyl ether
Concen: 201.99 ug/l
RT: 7.468 min Scan# 1906
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 63 Resp: 155042
Ion Ratio Lower Upper
63 100
106 24.4 20.3 30.5





#59

2-Hexanone

Concen: 236.84 ug/l

RT: 8.690 min Scan# 2286

Delta R.T. -0.004 min

Lab File: VU043403.D





Acq: 28 Apr 2021 22:14

Tgt Ion: 43 Resp: 525775

Ion Ratio Lower Upper

43 100

58 46.9 24.0 72.0



#60

Dibromochloromethane

Concen: 51.19 ug/l

RT: 8.816 min Scan# 2325

Delta R.T. -0.003 min

Lab File: VU043403.D




Acq: 28 Apr 2021 22:14


Tgt Ion:129 Resp: 67232

Ion Ratio Lower Upper

129 100

127 76.8 38.5 115.3









#61
1,2-Dibromoethane
Concen: 50.35 ug/l
RT: 8.931 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :

Tgt Ion:107 Resp: 64492
Ion Ratio Lower Upper
107 100
109 93.1 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 48.42 ug/l
RT: 10.639 min Scan# 2892
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 95 Resp: 72620
Ion Ratio Lower Upper
95 100
174 83.4 0.0 172.2
176 81.1 0.0 168.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 9.423 min Scan# 2514
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:117 Resp: 134686
Ion Ratio Lower Upper
117 100
82 57.2 45.8 68.6
119 31.6 25.8 38.8



#64
Tetrachloroethene
Concen: 58.19 ug/l
RT: 8.558 min Scan# 2245
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion:164 Resp: 56866
Ion Ratio Lower Upper
164 100
166 127.3 105.3 157.9
129 93.7 75.0 112.6
131 88.0 71.0 106.6









#65
Chlorobenzene
Concen: 49.53 ug/l
RT: 9.452 min Scan# 2523
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:112 Resp: 143047
Ion Ratio Lower Upper
112 100
114 30.8 25.0 37.4



#66
1,1,1,2-Tetrachloroethane
Concen: 51.88 ug/l
RT: 9.539 min Scan# 2550
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion:131 Resp: 57286
Ion Ratio Lower Upper
131 100
133 93.9 47.6 142.9
119 63.6 31.9 95.5





#67

Ethyl Benzene

Concen: 50.66 ug/l

RT: 9.578 min Scan# 2562

Delta R.T. -0.000 min

Lab File: VU043403.D





Acq: 28 Apr 2021 22:14

Tgt Ion: 91 Resp: 264066

Ion Ratio Lower Upper

91 100

106 29.7 24.1 36.1



#68

m/p-Xylenes

Concen: 100.92 ug/l

RT: 9.700 min Scan# 2600

Delta R.T. -0.000 min

Lab File: VU043403.D




Acq: 28 Apr 2021 22:14


Tgt Ion:106 Resp: 194647

Ion Ratio Lower Upper

106 100

91 214.3 169.9 254.9









#69
o-Xylene
Concen: 49.66 ug/l
RT: 10.108 min Scan# 2727
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:106 Resp: 93488
Ion Ratio Lower Upper
106 100
91 224.1 111.2 333.6



#70
Styrene
Concen: 51.61 ug/l
RT: 10.121 min Scan# 2731
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion:104 Resp: 163573
Ion Ratio Lower Upper
104 100
78 54.9 44.8 67.2
103 56.9 45.7 68.5









#71
Bromoform
Concen: 48.73 ug/l
RT: 10.298 min Scan# 2786
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:173 Resp: 55444
Ion Ratio Lower Upper
173 100
175 47.7 23.8 71.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 11.819 min Scan# 3259
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion:152 Resp: 77040
Ion Ratio Lower Upper
152 100
115 82.9 41.5 124.6
150 172.9 0.0 342.4





#73

Isopropylbenzene

Concen: 49.40 ug/l

RT: 10.491 min Scan# 2846

Delta R.T. -0.000 min





Lab File: VU043403.D

Acq: 28 Apr 2021 22:14

Tgt Ion:105 Resp: 255449

Ion Ratio Lower Upper

105	100		
120	25.5	13.1	39.3



#74

N-ethyl acetate

Concen: 45.67 ug/l

RT: 10.327 min Scan# 2795

Delta R.T. -0.000 min




Lab File: VU043403.D


Acq: 28 Apr 2021 22:14

Tgt Ion: 43 Resp: 151719

Ion Ratio Lower Upper

43	100		
70	34.1	28.3	42.5
55	22.3	17.2	25.8
61	19.6	17.3	25.9









#75
1,1,2,2-Tetrachloroethane
Concen: 48.91 ug/l
RT: 10.790 min Scan# 2939
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 83 Resp: 110127
Ion Ratio Lower Upper
83 100
131 9.6 5.1 15.2
85 64.1 31.7 95.1



#76
1,2,3-Trichloropropane
Concen: 54.28 ug/l
RT: 10.832 min Scan# 2952
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 75 Resp: 122201
Ion Ratio Lower Upper
75 100
77 29.8 20.2 60.5









#77
Bromobenzene
Concen: 48.35 ug/l
RT: 10.790 min Scan# 2939
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:156 Resp: 69642
Ion Ratio Lower Upper
156 100
77 160.8 79.3 238.1
158 97.0 48.3 144.9



#78
n-propylbenzene
Concen: 49.93 ug/l
RT: 10.912 min Scan# 2977
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 91 Resp: 309117
Ion Ratio Lower Upper
91 100
120 21.8 11.2 33.6









#79
2-Chlorotoluene
Concen: 48.91 ug/l
RT: 10.992 min Scan# 3002
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 91 Resp: 184002
Ion Ratio Lower Upper
91 100
126 33.4 16.7 50.0



#80
1,3,5-Trimethylbenzene
Concen: 49.38 ug/l
RT: 11.095 min Scan# 3034
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14


Tgt Ion: 105 Resp: 225398
Ion Ratio Lower Upper
105 100
120 46.8 23.8 71.4








#81
trans-1,4-Dichloro-2-butene
Concen: 38.26 ug/l
RT: 10.552 min Scan# 2865
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Instrument :
MSVOA_U
ClientSampled :







Tgt Ion	Ratio	Lower	Upper
75	100		
53	119.4	85.6	128.4
89	48.2	35.7	53.5



#82
4-Chlorotoluene
Concen: 48.94 ug/l
RT: 11.102 min Scan# 3036
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion	Ratio	Lower	Upper
91	100		
126	28.8	14.6	44.0









#83
tert-Butylbenzene
Concen: 47.47 ug/l
RT: 11.426 min Scan# 3137
Delta R.T. -0.001 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:119 Resp: 213032
Ion Ratio Lower Upper
119 100
91 62.1 30.3 90.9
134 22.0 11.4 34.1



#84
1,2,4-Trimethylbenzene
Concen: 49.10 ug/l
RT: 11.475 min Scan# 3152
Delta R.T. -0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion:105 Resp: 230263
Ion Ratio Lower Upper
105 100
120 43.9 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 47.91 ug/l
 RT: 11.651 min Scan# 3207
 Delta R.T. -0.001 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14




Instrument :
 MSVOA_U
 ClientSampled :


Tgt Ion:105 Resp: 251933
 Ion Ratio Lower Upper
 105 100
 134 19.3 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 47.73 ug/l
 RT: 11.799 min Scan# 3253
 Delta R.T. 0.000 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14

Tgt Ion:119 Resp: 231071
 Ion Ratio Lower Upper
 119 100
 134 25.8 13.2 39.6
 91 25.4 12.8 38.3





#87

1,3-Dichlorobenzene

Concen: 46.99 ug/l

RT: 11.751 min Scan# 3238

Delta R.T. 0.000 min

Lab File: VU043403.D

Acq: 28 Apr 2021 22:14





Tgt Ion:146 Resp: 122706

Ion Ratio Lower Upper

146 100

111 40.2 19.9 59.6

148 65.7 32.0 96.2



#88

1,4-Dichlorobenzene

Concen: 46.50 ug/l

RT: 11.841 min Scan# 3266

Delta R.T. 0.000 min

Lab File: VU043403.D

Acq: 28 Apr 2021 22:14




Tgt Ion:146 Resp: 126643


Ion Ratio Lower Upper

146 100

111 38.7 19.7 59.0

148 64.5 32.8 98.3









#89
n-Butylbenzene
Concen: 45.64 ug/l
RT: 12.214 min Scan# 3382
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 91 Resp: 206958
Ion Ratio Lower Upper
91 100
92 50.9 26.2 78.6
134 24.0 12.7 38.1



#90
Hexachloroethane
Concen: 39.49 ug/l
RT: 12.484 min Scan# 3466
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 117 Resp: 35837
Ion Ratio Lower Upper
117 100
201 81.0 43.5 130.7









#91
1,2-Dichlorobenzene
Concen: 48.08 ug/l
RT: 12.217 min Scan# 3383
Delta R.T. -0.004 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:146 Resp: 128290
Ion Ratio Lower Upper
146 100
111 41.0 20.8 62.4
148 64.9 31.9 95.7



#92
1,2-Dibromo-3-Chloropropane
Concen: 43.47 ug/l
RT: 13.002 min Scan# 3627
Delta R.T. -0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion: 75 Resp: 27708
Ion Ratio Lower Upper
75 100
155 76.6 41.7 125.1
157 98.6 53.9 161.7









#93
 1,2,4-Trichlorobenzene
 Concen: 46.84 ug/l
 RT: 13.848 min Scan# 3890
 Delta R.T. -0.000 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14




Instrument :
 MSVOA_U
 ClientSampled :


Tgt Ion:180 Resp: 82466
 Ion Ratio Lower Upper
 180 100
 182 96.9 46.9 140.8
 145 29.8 14.4 43.4



#94
 Hexachlorobutadiene
 Concen: 42.89 ug/l
 RT: 14.028 min Scan# 3946
 Delta R.T. -0.000 min
 Lab File: VU043403.D
 Acq: 28 Apr 2021 22:14


Tgt Ion:225 Resp: 37932
 Ion Ratio Lower Upper
 225 100
 223 64.8 32.3 96.9
 227 65.8 31.8 95.4








#95
Naphthalene
Concen: 49.82 ug/l
RT: 14.095 min Scan# 3967
Delta R.T. 0.003 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Instrument :
MSVOA_U
ClientSampled :



Tgt Ion:128 Resp: 277580
Ion Ratio Lower Upper
128 100
127 13.7 10.8 16.2
129 10.8 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 50.32 ug/l
RT: 14.336 min Scan# 4042
Delta R.T. 0.000 min
Lab File: VU043403.D
Acq: 28 Apr 2021 22:14

Tgt Ion:180 Resp: 86088
Ion Ratio Lower Upper
180 100
182 96.3 48.0 144.0
145 31.6 15.5 46.5

