

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042921\
 Data File : VU043411.D
 Acq On : 29 Apr 2021 01:42
 Operator : SY/MD
 Sample : M2180-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-4

Quant Time: Apr 29 09:47:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	101227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	167759	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	164809	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	76965	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	97535	57.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.64%	
35) Dibromofluoromethane	5.298	113	67123	53.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.50%	
50) Toluene-d8	7.906	98	229878	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
62) 4-Bromofluorobenzene	10.639	95	80554	45.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.20%	
Target Compounds						
16) Acetone	2.630	43	2101	1.80	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042921\
Data File : VU043411.D
Acq On : 29 Apr 2021 01:42
Operator : SY/MD
Sample : M2180-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 63 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-4

Quant Time: Apr 29 09:47:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

