

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042922\
 Data File : VU048349.D
 Acq On : 29 Apr 2022 10:39
 Operator : SY/MD
 Sample : VU0429MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK107

Quant Time: Apr 30 04:39:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 30 04:38:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	386817	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	378890	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	170240	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	142037	39.931	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.860%
7) Chloroethane-d5	1.906	69	117193	46.937	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.880%
11) 1,1-Dichloroethene-d2	2.565	63	195003	34.214	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.420%
21) 2-Butanone-d5	4.629	46	302291	99.640	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.640%
24) Chloroform-d	5.063	84	299062	46.352	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.700%
26) 1,2-Dichloroethane-d4	5.703	65	193119	50.448	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.900%
32) Benzene-d6	5.729	84	622851	52.123	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.240%
36) 1,2-Dichloropropane-d6	6.690	67	203698	51.561	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.120%
41) Toluene-d8	7.899	98	550932	50.863	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.720%
43) trans-1,3-Dichloroprop...	8.179	79	86457	49.596	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.200%
47) 2-Hexanone-d5	8.639	63	219904	103.172	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.170%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	324109	46.778	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.560%
66) 1,2-Dichlorobenzene-d4	12.195	152	208887	54.545	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.100%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042922\
Data File : VU048349.D
Acq On : 29 Apr 2022 10:39
Operator : SY/MD
Sample : VU0429MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK107

Quant Time: Apr 30 04:39:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 30 04:38:24 2022
Response via : Initial Calibration

