

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042922\
 Data File : VU048350.D
 Acq On : 29 Apr 2022 11:04
 Operator : SY/MD
 Sample : VU0429WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK108

Quant Time: Apr 30 04:39:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 30 04:38:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	348461	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	353954	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	176568	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	126729	39.549	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.100%
7) Chloroethane-d5	1.896	69	108342	48.168	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.340%
11) 1,1-Dichloroethene-d2	2.565	63	173690	33.829	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.660%
21) 2-Butanone-d5	4.620	46	269682	98.676	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.680%
24) Chloroform-d	5.063	84	271493	46.711	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.420%
26) 1,2-Dichloroethane-d4	5.700	65	176420	51.158	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.320%
32) Benzene-d6	5.726	84	555605	49.771	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.540%
36) 1,2-Dichloropropane-d6	6.690	67	185381	50.231	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.460%
41) Toluene-d8	7.899	98	490971	48.521	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.040%
43) trans-1,3-Dichloroprop...	8.179	79	79263	48.673	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.340%
47) 2-Hexanone-d5	8.635	63	209260	105.094	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.090%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	325137	50.232	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.460%
66) 1,2-Dichlorobenzene-d4	12.195	152	212612	53.528	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.060%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042922\
Data File : VU048350.D
Acq On : 29 Apr 2022 11:04
Operator : SY/MD
Sample : VU0429WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK108

Quant Time: Apr 30 04:39:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042522WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 30 04:38:24 2022
Response via : Initial Calibration

